

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011521\
 Data File : PS013781.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jan 2021 12:34
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 1/18/2021 12:00:20 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 15 13:42:47 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010421.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 05 01:51:07 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.206	6.908	778.1E6	300.9E6	814.308	748.291
Target Compounds							
1) T	Dalapon	3.029	2.623	792.0E6	342.9E6	728.148	659.421
2) T	3,5-DICHL...	6.423	6.064	1013.4E6	440.1E6	766.339	698.596
3) T	4-Nitroph...	7.041	6.522	355.1E6	176.4E6	746.798	695.524
5) T	DICAMBA	7.396	7.075	3054.8E6	1235.1E6	793.239	719.187
6) T	MCPD	7.563	7.169	225.0E6	100.7E6	78.553	71.272
7) T	MCPA	7.712	7.384	318.4E6	141.6E6	77.466	70.060
8) T	DICHLORPROP	8.086	7.699	758.3E6	315.9E6	783.340	709.424
9) T	2,4-D	8.318	7.997	811.3E6	390.8E6	765.868m	704.875
10) T	Pentachlo...	8.625	8.420	11034.7E6	4154.1E6	778.887	718.620
11) T	2,4,5-TP ...	9.174	8.782	4436.6E6	1732.7E6	796.727	725.348
12) T	2,4,5-T	9.460	9.163	4458.5E6	1668.4E6	788.471	728.771
13) T	2,4-DB	10.014	9.670	732.9E6	220.7E6	750.002	725.771
14) T	DINOSEB	11.166	9.997	3423.4E6	1214.2E6	775.963	725.649
15) T	Picloram	11.002	10.993	5325.5E6	2256.0E6	737.333	649.094
16) T	DCPA	11.431	10.924	5635.1E6	2254.1E6	797.584	728.295

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011521\
Data File : PS013781.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jan 2021 12:34
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
1/18/2021 12:00:20 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 15 13:42:47 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010421.M
Quant Title : 8080.M
QLast Update : Tue Jan 05 01:51:07 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

