

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011719\
 Data File : PS002842.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jan 2019 02:27
 Operator : EI\SJ
 Sample : K1167-02
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 JDHS-SOUTH-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 04:15:47 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011719.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 17 13:42:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.295	10.788	220.6E6	219.7E6	611.656	544.675
Target Compounds							
1) T	Dalapon	2.806	2.287	48329861	113.3E6	97.479	225.361 #
2) T	3,5-DICHL...	9.753	0.000	11211075	0	20.882	N.D. #
5) T	DICAMBA	0.000	11.028	0	10051202	N.D.	6.904 #
6) T	MCP	11.959	11.349	50035269	3169615	48.404	3.342 #
7) T	MCPA	12.145	11.744	29533615	75434049	17.592	43.161 #
8) T	DICHLORPROP	0.000	12.193	0	10672491	N.D.	28.396 #
10) T	Pentachlo...	13.531	13.202	10104055	4494470	1.840	0.809 #
11) T	2,4,5-TP ...	14.423	13.838	1553740	9455751	0.746	4.363 #
12) T	2,4,5-T	14.823	14.439	14232855	1109237	6.707	0.657 #
13) T	2,4-DB	15.569	15.024	12531990	5936362	48.598	21.674 #
14) T	DINOSEB	16.958	15.448	18324853	22983037	10.989	13.915 #
15) T	Picloram	16.746	16.733	8899661	28999302	3.303	9.893 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011719\
Data File : PS002842.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jan 2019 02:27
Operator : EI\SJ
Sample : K1167-02
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
JDHS-SOUTH-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 04:15:47 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011719.M
Quant Title : 8080.M
QLast Update : Thu Jan 17 13:42:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.806 min
Delta R.T.: -0.046 min
Response: 48329861
Conc: 97.48 ng/ml

Instrument :
ECD_S
ClientSampleId :
JDHS-SOUTH-1




#1 Dalapon

R.T.: 2.287 min
Delta R.T.: -0.032 min
Response: 113325547
Conc: 225.36 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 9.753 min
Delta R.T.: 0.027 min
Response: 11211075
Conc: 20.88 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 0.000 min
Exp R.T.: 8.949 min
Response: 0
Conc: N.D.



#4 2,4-DCAA


R.T.: 11.295 min
Delta R.T.: -0.004 min
Response: 220615483
Conc: 611.66 ng/ml

Instrument :
ECD_S
ClientSampleId :
JDHS-SOUTH-1




#4 2,4-DCAA

R.T.: 10.788 min
Delta R.T.: -0.002 min
Response: 219677838
Conc: 544.68 ng/ml




#5 DICAMBA

R.T.: 0.000 min
Exp R.T. : 11.607 min
Response: 0
Conc: N.D.



#5 DICAMBA


R.T.: 11.028 min
Delta R.T.: -0.044 min
Response: 10051202
Conc: 6.90 ng/ml



#6 MCPP


R.T.: 11.959 min
Delta R.T.: 0.001 min
Response: 50035269
Conc: 48.40 ug/ml

Instrument :
ECD_S
ClientSampleId :
JDHS-SOUTH-1




#6 MCPP

R.T.: 11.349 min
Delta R.T.: 0.038 min
Response: 3169615
Conc: 3.34 ug/ml




#7 MCPA

R.T.: 12.145 min
Delta R.T.: -0.065 min
Response: 29533615
Conc: 17.59 ug/ml



#7 MCPA


R.T.: 11.744 min
Delta R.T.: 0.056 min
Response: 75434049
Conc: 43.16 ug/ml



#8 DICHLORPROP


R.T.: 0.000 min
Exp R.T.: 12.797 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
JDHS-SOUTH-1




#8 DICHLORPROP

R.T.: 12.193 min
Delta R.T.: -0.004 min
Response: 10672491
Conc: 28.40 ng/ml




#10 Pentachlorophenol

R.T.: 13.531 min
Delta R.T.: 0.004 min
Response: 10104055
Conc: 1.84 ng/ml



#10 Pentachlorophenol


R.T.: 13.202 min
Delta R.T.: -0.002 min
Response: 4494470
Conc: 0.81 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 14.423 min
Delta R.T.: 0.020 min
Response: 1553740
Conc: 0.75 ng/ml

Instrument :
ECD_S
ClientSampleId :
JDHS-SOUTH-1




#11 2,4,5-TP (SILVEX)

R.T.: 13.838 min
Delta R.T.: 0.009 min
Response: 9455751
Conc: 4.36 ng/ml




#12 2,4,5-T

R.T.: 14.823 min
Delta R.T.: 0.023 min
Response: 14232855
Conc: 6.71 ng/ml



#12 2,4,5-T


R.T.: 14.439 min
Delta R.T.: 0.048 min
Response: 1109237
Conc: 0.66 ng/ml



#13 2,4-DB


R.T.: 15.569 min
Delta R.T.: 0.023 min
Response: 12531990
Conc: 48.60 ng/ml

Instrument :
ECD_S
ClientSampleId :
JDHS-SOUTH-1




#13 2,4-DB

R.T.: 15.024 min
Delta R.T.: -0.019 min
Response: 5936362
Conc: 21.67 ng/ml




#14 DINOSEB

R.T.: 16.958 min
Delta R.T.: -0.064 min
Response: 18324853
Conc: 10.99 ng/ml



#14 DINOSEB


R.T.: 15.448 min
Delta R.T.: -0.021 min
Response: 22983037
Conc: 13.91 ng/ml



#15 Picloram

R.T.: 16.746 min
Delta R.T.: -0.043 min
Response: 8899661
Conc: 3.30 ng/ml

Instrument :
ECD_S
ClientSampleId :
JDHS-SOUTH-1



#15 Picloram

R.T.: 16.733 min
Delta R.T.: 0.036 min
Response: 28999302
Conc: 9.89 ng/ml