

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011724\
 Data File : PS025571.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jan 2024 22:28
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 17 23:03:27 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011124.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 11 16:45:24 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	6.351	6.760	754.7E6	402.6E6	506.679	524.387
Target Compounds						
1) T Dalapon	0.000	2.558	0	3225735	N.D.	2.746
3) T 4-Nitroph...	0.000	6.367	0	1221161	N.D.	2.565
7) T MCPA	0.000	7.153f	0	11003063	N.D.	3.176
8) T DICHLORPROP	0.000	7.574f	0	5784886	N.D.	6.692
15) T Picloram	0.000	10.799f	0	316917	N.D.	<MDL



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011724\
Data File : PS025571.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jan 2024 22:28
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 17 23:03:27 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011124.M
Quant Title : 8080.M
QLast Update : Thu Jan 11 16:45:24 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 0.000 min
Exp R.T.: 2.474 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
I.BLK




#1 Dalapon

R.T.: 2.558 min
Delta R.T.: 0.029 min
Response: 3225735
Conc: 2.75 ng/ml




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T.: 6.191 min
Response: 0
Conc: N.D.



#3 4-Nitrophenol


R.T.: 6.367 min
Delta R.T.: -0.012 min
Response: 1221161
Conc: 2.56 ng/ml



#4 2,4-DCAA


R.T.: 6.351 min
Delta R.T.: 0.001 min
Response: 754680025
Conc: 506.68 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK




#4 2,4-DCAA

R.T.: 6.760 min
Delta R.T.: -0.001 min
Response: 402624458
Conc: 524.39 ng/ml




#7 MCPA

R.T.: 0.000 min
Exp R.T.: 6.762 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 7.153 min
Delta R.T.: -0.068 min
Response: 11003063
Conc: 3.18 ug/ml



#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T.: 7.077 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
I.BLK



#8 DICHLORPROP

R.T.: 7.574 min
Delta R.T.: 0.042 min
Response: 5784886
Conc: 6.69 ng/ml