

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011725\
 Data File : PS028932.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jan 2025 20:05
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 03:24:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 14 12:25:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.196	7.675	2030.5E6	833.6E6	729.347	747.035
Target Compounds							
1) T	Dalapon	2.616	2.667	2262.6E6	1380.2E6	758.804	676.533
2) T	3,5-DICHL...	6.373	6.641	2763.7E6	1148.6E6	691.479	695.028
3) T	4-Nitroph...	6.994	7.205	1217.1E6	619.3E6	686.816	696.071
5) T	DICAMBA	7.382	7.872	8427.3E6	4084.8E6	710.476	733.486
6) T	MCPD	7.564	7.977	526.3E6	211.4E6	77.208	70.267
7) T	MCPA	7.713	8.219	717.0E6	290.2E6	72.786	68.314
8) T	DICHLORPROP	8.086	8.584	2175.7E6	1000.6E6	686.644	711.839
9) T	2,4-D	8.315	8.911	2346.8E6	1058.7E6	694.424	706.036
10) T	Pentachlo...	8.612	9.434	34273.8E6	17155.0E6	710.529	740.524
11) T	2,4,5-TP ...	9.187	9.811	13435.0E6	6962.7E6	702.200	739.186
12) T	2,4,5-T	9.478	10.227	13564.3E6	6612.8E6	706.585	734.020
13) T	2,4-DB	10.049	10.792	2516.9E6	716.7E6	709.518	719.709
14) T	DINOSEB	11.252	11.169	11266.2E6	4592.1E6	680.846	715.580
15) T	Picloram	11.061	12.251	21896.3E6	9898.0E6	693.972	737.578
16) T	DCPA	11.546	12.207	20261.4E6	8598.1E6	706.438	757.388

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011725\
Data File : PS028932.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jan 2025 20:05
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 03:24:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

