

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011819\
 Data File : PS002852.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jan 2019 08:32
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 1/21/2019 11:20:55 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 11:19:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011719.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 17 13:42:13 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.294	10.787	286.9E6	313.2E6	795.528	776.641
Target Compounds							
1) T	Dalapon	2.851	2.319	348.2E6	350.6E6	702.212	697.267
2) T	3,5-DICHL...	9.722	8.946	391.4E6	359.4E6	728.982	723.619m
3) T	4-Nitroph...	10.879	9.935	155.2E6	139.2E6	735.281	693.502
5) T	DICAMBA	11.601	11.068	1150.1E6	1108.9E6	767.892	761.740
6) T	MCPD	11.945	11.300	79188858	72899164	76.607m	76.860
7) T	MCPA	12.196	11.676	124.8E6	119.7E6	74.350	68.486
8) T	DICHLORPROP	12.792	12.193	282.5E6	282.4E6	738.197	751.407
9) T	2,4-D	13.151	12.665	303.1E6	334.5E6	745.968	762.624
10) T	Pentachlo...	13.520	13.201	4259.7E6	4276.3E6	775.602	769.452
11) T	2,4,5-TP ...	14.397	13.826	1612.5E6	1679.9E6	774.365	775.035
12) T	2,4,5-T	14.794	14.388	1649.3E6	1258.3E6	777.172	744.943m
13) T	2,4-DB	15.541	15.040	204.5E6	188.6E6	792.890	688.502m
14) T	DINOSEB	17.012	15.466	1113.9E6	1175.6E6	667.968m	711.761m
15) T	Picloram	16.782	16.695	2109.7E6	1991.9E6	783.068	679.541m
16) T	DCPA	17.358	16.641	1998.8E6	2216.2E6	745.794m	771.685m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011819\
Data File : PS002852.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jan 2019 08:32
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
1/21/2019 11:20:55 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 11:19:32 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011719.M
Quant Title : 8080.M
QLast Update : Thu Jan 17 13:42:13 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

