

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011821\
 Data File : PS013796.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jan 2021 16:26
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 1/19/2021 10:03:32 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 16:54:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010421.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 05 01:51:07 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.206	6.908	807.5E6	309.9E6	845.032	770.563
Target Compounds							
1) T	Dalapon	3.029	2.623	829.9E6	357.9E6	763.082	688.363
2) T	3,5-DICHL...	6.423	6.063	1035.8E6	449.0E6	783.287	712.754
3) T	4-Nitroph...	7.040	6.521	370.3E6	182.4E6	778.955	719.277
5) T	DICAMBA	7.395	7.074	3073.7E6	1261.7E6	798.136m	734.705
6) T	MCPD	7.563	7.170	230.4E6	102.5E6	80.435	72.567
7) T	MCPA	7.712	7.384	325.8E6	144.5E6	79.265	71.458
8) T	DICHLORPROP	8.087	7.699	773.3E6	323.1E6	798.831	725.463
9) T	2,4-D	8.318	7.996	812.3E6	401.5E6	766.821m	724.221
10) T	Pentachlo...	8.625	8.421	11261.9E6	4231.9E6	794.926	732.075
11) T	2,4,5-TP ...	9.173	8.782	4514.9E6	1761.3E6	810.788	737.324
12) T	2,4,5-T	9.460	9.162	4564.3E6	1716.0E6	807.181	749.579
13) T	2,4-DB	10.013	9.669	781.3E6	232.2E6	799.588	763.302
14) T	DINOSEB	11.166	9.997	3457.3E6	1240.1E6	783.640	741.100
15) T	Picloram	10.998	10.990	5821.5E6	2427.9E6	806.009	698.566
16) T	DCPA	11.432	10.924	5692.3E6	2284.0E6	805.683	737.969

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011821\
Data File : PS013796.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jan 2021 16:26
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
1/19/2021 10:03:32 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 16:54:50 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010421.M
Quant Title : 8080.M
QLast Update : Tue Jan 05 01:51:07 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

