

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011821\
 Data File : PS013802.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jan 2021 18:44
 Operator : DD\AJ
 Sample : M1141-21MSD
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 T872-T871MSD

Manual Integrations
 APPROVED

Ankita
 1/19/2021 10:03:38 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 19 01:47:19 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010421.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 05 01:51:07 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.207	6.909	568.2E6	248.8E6	594.562	618.620
Target Compounds							
1) T	Dalapon	3.025	2.621	621.8E6	318.3E6	571.696	612.116m
2) T	3,5-DICHL...	6.422	6.064	801.7E6	310.6E6	606.271	493.060
5) T	DICAMBA	7.395	7.075	2223.8E6	913.7E6	577.464	532.044
6) T	MCPD	7.562	7.168	175.6E6	74922881	61.297	53.019
7) T	MCPA	7.711	7.383	252.5E6	120.4E6	61.429	59.578
8) T	DICHLORPROP	8.086	7.700	578.2E6	242.1E6	597.339	543.583
9) T	2,4-D	8.318	7.997	691.5E6	326.0E6	652.790	587.959
10) T	Pentachlo...	8.625	8.421	8776.2E6	3299.5E6	619.473	570.775
11) T	2,4,5-TP ...	9.173	8.782	3582.8E6	1452.5E6	643.407	608.075
12) T	2,4,5-T	9.460	9.163	3572.0E6	1340.4E6	631.699	585.517
13) T	2,4-DB	10.012	9.669	458.6E6	172.3E6	469.281	566.579
14) T	DINOSEB	11.166	9.998	2509.2E6	908.3E6	568.750	542.816
15) T	Picloram	10.999	10.991	3684.9E6	1619.3E6	510.194	465.921
16) T	DCPA	11.432	10.924	4281.5E6	1762.6E6	605.999	569.481

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS011821\
Data File : PS013802.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jan 2021 18:44
Operator : DD\AJ
Sample : M1141-21MSD
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
T872-T871MSD

Manual Integrations
APPROVED

Ankita
1/19/2021 10:03:38 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 19 01:47:19 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010421.M
Quant Title : 8080.M
QLast Update : Tue Jan 05 01:51:07 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

