

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012020\
 Data File : PS008463.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jan 2020 08:56
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 1/22/2020 9:38:54 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 22 08:56:52 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010820.M
 Quant Title : 8080.M
 QLast Update : Wed Jan 08 11:00:59 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.271	5.966	1428.4E6	708.4E6	808.996	747.644
Target Compounds							
1) T	Dalapon	2.536	2.227	1823.5E6	861.7E6	711.251	679.708
2) T	3,5-DICHL...	5.491	5.123	1820.6E6	975.6E6	760.102	698.283
3) T	4-Nitroph...	6.106	5.573	695.9E6	393.7E6	714.547	663.074
5) T	DICAMBA	6.465	6.137	5392.5E6	2864.1E6	784.623	716.208
6) T	MCPD	6.633	6.232	411.2E6	217.2E6	80.668	71.438
7) T	MCPA	6.781	6.448	584.0E6	303.2E6	79.923	71.012m
8) T	DICHLORPROP	7.154	6.760	1280.7E6	680.2E6	780.630	704.023
9) T	2,4-D	7.381	7.054	1381.2E6	817.9E6	759.295	697.682
10) T	Pentachlo...	7.697	7.489	18834.9E6	9589.0E6	774.755	710.015
11) T	2,4,5-TP ...	8.244	7.848	7246.8E6	3719.0E6	775.990	711.551
12) T	2,4,5-T	8.527	8.227	7017.7E6	3728.5E6	760.601	705.073
13) T	2,4-DB	9.080	8.735	1010.9E6	495.2E6	729.237	699.520
14) T	DINOSEB	10.247	9.071	5315.2E6	2660.8E6	740.809	704.333
15) T	Picloram	10.059	10.052	9171.3E6	5345.7E6	734.782	719.544
16) T	DCPA	10.511	10.003	8897.7E6	4708.7E6	749.730	722.556

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012020\
Data File : PS008463.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jan 2020 08:56
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
1/22/2020 9:38:54 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 22 08:56:52 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010820.M
Quant Title : 8080.M
QLast Update : Wed Jan 08 11:00:59 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

