

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012023\
 Data File : PS021826.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jan 2023 14:46
 Operator : AR\AJ
 Sample : PB150359BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB150359BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 20 21:09:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010923.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 10 10:34:52 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.375	6.809	1414.1E6	965.5E6	508.013	537.390
Target Compounds							
1) T	Dalapon	2.483	2.560	1192.8E6	1121.9E6	421.637	435.830
2) T	3,5-DICHL...	5.748	5.991	1837.8E6	1368.5E6	475.101	496.570
3) T	4-Nitroph...	6.217	6.427	774.8E6	569.2E6	531.585	519.192
5) T	DICAMBA	6.516	6.973	5904.0E6	4283.9E6	482.860	503.826
6) T	MCPD	6.671	7.069	441.8E6	320.0E6	46.107	48.253
7) T	MCPA	6.796	7.281	600.4E6	438.5E6	46.482	48.345
8) T	DICHLORPROP	7.113	7.598	1545.5E6	1071.4E6	479.186	502.182
9) T	2,4-D	7.318	7.891	1841.1E6	1284.3E6	474.047	508.328
10) T	Pentachlo...	7.543	8.316	21975.9E6	15191.8E6	527.578	529.359
11) T	2,4,5-TP ...	8.072	8.678	9147.8E6	6154.4E6	496.459	517.598
12) T	2,4,5-T	8.333	9.055	9428.7E6	5927.5E6	489.150	530.667
13) T	2,4-DB	8.844	9.562	1584.0E6	792.7E6	500.415	529.050
14) T	DINOSEB	9.894	9.888	7786.0E6	4251.6E6	474.362	511.574
15) T	Picloram	9.735	10.862	13069.2E6	8175.5E6	460.711	461.669
16) T	DCPA	10.162	10.815	12355.4E6	7344.5E6	500.143	526.874

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012023\
Data File : PS021826.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jan 2023 14:46
Operator : AR\AJ
Sample : PB150359BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB150359BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 20 21:09:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010923.M
Quant Title : 8080.M
QLast Update : Tue Jan 10 10:34:52 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

