

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\
 Method File : PS012219.M
 Title : 8080.M
 Last Update : Wed Jan 23 08:15:42 2019
 Response Via : Initial Calibration

Calibration Files

750 =PS002922.D 1000 =PS002923.D 1500 =PS002924.D
 500 =PS002921.D 200 =PS002920.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	4.770	4.907	4.907	4.831	4.903	4.863 E5	1.26
2) T	3,5-DICHLOROBENZOI...	5.224	5.286	5.169	5.121	5.129	5.186 E5	1.34
3) T	4-Nitrophenol	1.986	1.878	1.826	1.806	1.618	1.823 E5	7.36
4) S	2,4-DCAA	3.344	3.437	3.387	3.340	3.357	3.373 E5	1.19
5) T	DICAMBA	1.512	1.548	1.541	1.480	1.473	1.511 E6	2.25
6) T	MCPP	1.044	1.113	1.129	0.997	0.859	1.029 E6	10.55
7) T	MCPA	1.606	1.685	1.694	1.539	1.401	1.585 E6	7.62
8) T	DICHLORPROP	3.678	3.771	3.697	3.681	3.748	3.715 E5	1.13
9) T	2,4-D	3.871	4.021	4.056	3.842	3.671	3.892 E5	3.97
10) T	Pentachlorophenol	5.795	5.849	5.590	5.695	5.717	5.729 E6	1.73
11) T	2,4,5-TP (SILVEX)	2.144	2.168	2.133	2.098	2.089	2.126 E6	1.54
12) T	2,4,5-T	2.090	2.109	2.071	2.009	1.930	2.042 E6	3.58
13) T	2,4-DB	2.412	2.500	2.548	2.389	2.234	2.417 E5	5.00
14) T	DINOSEB	1.565	1.612	1.589	1.522	1.528	1.563 E6	2.47
15) T	Picloram	2.882	3.038	2.996	2.794	2.556	2.853 E6	6.73
16) T	DCPA	2.668	2.719	2.648	2.635	2.612	2.656 E6	1.53

Signal #2 Calibration Files

750 =PS002922.D 1000 =PS002923.D 1500 =PS002924.D
 500 =PS002921.D 200 =PS002920.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	4.667	4.734	4.777	4.591	4.883	4.730 E5	2.34
2) T	3,5-DICHLOROBENZOI...	4.910	4.982	4.930	4.734	4.904	4.892 E5	1.92
3) T	4-Nitrophenol	1.854	1.878	1.860	1.836	1.800	1.846 E5	1.61
4) S	2,4-DCAA	3.884	4.016	3.974	3.877	3.860	3.922 E5	1.75
5) T	DICAMBA	1.471	1.521	1.515	1.449	1.427	1.477 E6	2.79
6) T	MCPP	0.970	1.015	1.031	0.925	0.807	0.950 E6	9.48
7) T	MCPA	1.522	1.605	1.627	1.419	1.251	1.485 E6	10.39
8) T	DICHLORPROP	3.727	3.769	3.829	3.698	3.775	3.760 E5	1.33
9) T	2,4-D	4.402	4.548	4.555	4.338	4.335	4.435 E5	2.46
10) T	Pentachlorophenol	5.804	5.830	5.561	5.735	5.796	5.745 E6	1.89
11) T	2,4,5-TP (SILVEX)	2.212	2.253	2.230	2.180	2.162	2.207 E6	1.67
12) T	2,4,5-T	1.743	1.730	1.802	1.748	1.644	1.734 E6	3.27
13) T	2,4-DB	2.550	2.672	2.715	2.495	2.410	2.568 E5	4.88
14) T	DINOSEB	1.669	1.717	1.691	1.622	1.636	1.667 E6	2.34
15) T	Picloram	3.007	3.030	3.078	2.900	2.728	2.949 E6	4.73
16) T	DCPA	2.968	2.966	2.866	2.936	2.960	2.939 E6	1.46

(#) = Out of Range