

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012219\
 Data File : PS002941.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jan 2019 22:55
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 1/23/2019 1:08:52 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 23 08:21:26 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012219.M
 Quant Title : 8080.M
 QLast Update : Wed Jan 23 08:15:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.281	10.780	259.6E6	302.4E6	769.566m	770.896m
Target Compounds							
1) T	Dalapon	2.838	2.307	331.6E6	333.0E6	681.825m	703.983m
2) T	3,5-DICHL...	9.709	8.939	366.9E6	353.4E6	707.466m	722.476m
3) T	4-Nitroph...	10.866	9.927	125.8E6	131.8E6	689.925m	713.867m
5) T	DICAMBA	11.588	11.061	1090.4E6	1079.2E6	721.713m	730.808m
6) T	MCPD	11.934	11.295	80402832	71419665	78.173m	75.208m
7) T	MCPA	12.184	11.670	119.6E6	110.3E6	75.496m	74.292m
8) T	DICHLORPROP	12.779	12.187	268.3E6	273.2E6	722.308m	726.746m
9) T	2,4-D	13.137	12.659	287.8E6	324.6E6	739.557m	731.940m
10) T	Pentachlo...	13.508	13.194	4317.3E6	4303.0E6	753.531m	748.945m
11) T	2,4,5-TP ...	14.384	13.821	1561.9E6	1634.5E6	734.518m	740.450m
12) T	2,4,5-T	14.781	14.384	1508.0E6	1280.4E6	738.609m	738.613m
13) T	2,4-DB	15.528	15.034	185.0E6	193.3E6	765.484m	752.482m
14) T	DINOSEB	17.000	15.460	1157.9E6	1227.7E6	740.713m	736.413m
15) T	Picloram	16.769	16.689	2121.4E6	2251.8E6	743.548m	763.714m
16) T	DCPA	17.346	16.636	2024.4E6	2225.6E6	762.085m	757.242m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012219\
Data File : PS002941.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jan 2019 22:55
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
1/23/2019 1:08:52 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 23 08:21:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012219.M
Quant Title : 8080.M
QLast Update : Wed Jan 23 08:15:42 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

