

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012220\
 Data File : PS008486.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2020 10:20
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 22 00:52:07 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010820.M
 Quant Title : 8080.M
 QLast Update : Wed Jan 08 11:00:59 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.285	5.981	1381.6E6	712.8E6	782.461	752.289
Target Compounds							
1) T	Dalapon	2.551	2.241	1772.7E6	829.9E6	691.437	654.610
2) T	3,5-DICHL...	5.505	5.138	1735.6E6	982.4E6	724.637	703.102
3) T	4-Nitroph...	6.120	5.588	654.6E6	397.7E6	672.048	669.784
5) T	DICAMBA	6.479	6.151	5140.9E6	2883.9E6	748.021	721.141
6) T	MCPD	6.647	6.247	394.4E6	219.3E6	77.391	72.126
7) T	MCPA	6.796	6.463	551.4E6	304.7E6	75.460	71.357
8) T	DICHLORPROP	7.168	6.775	1213.2E6	685.6E6	739.489	709.612
9) T	2,4-D	7.395	7.068	1321.9E6	829.8E6	726.651	707.876
10) T	Pentachlo...	7.711	7.504	17920.6E6	9661.4E6	737.145	715.378
11) T	2,4,5-TP ...	8.257	7.863	6939.8E6	3789.5E6	743.118	725.030
12) T	2,4,5-T	8.541	8.242	6733.1E6	3821.7E6	729.755	722.701
13) T	2,4-DB	9.093	8.750	959.4E6	504.9E6	692.065	713.185
14) T	DINOSEB	10.261	9.085	5103.5E6	2710.8E6	711.300	717.566
15) T	Picloram	10.072	10.065	8335.6E6	5077.2E6	667.826	683.405
16) T	DCPA	10.524	10.017	8514.6E6	4779.3E6	717.456	733.400

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012220\
Data File : PS008486.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2020 10:20
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 22 00:52:07 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010820.M
Quant Title : 8080.M
QLast Update : Wed Jan 08 11:00:59 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

