

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012220\
 Data File : PS008500.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2020 20:00
 Operator : DD\AJ
 Sample : L1004-12MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 BU-02-101720MS

Manual Integrations
 APPROVED

Ankita
 1/22/2020 10:57:44 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 22 00:55:23 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010820.M
 Quant Title : 8080.M
 QLast Update : Wed Jan 08 11:00:59 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.272	5.965	543.0E6	331.3E6	307.545	349.686
Target Compounds							
1) T	Dalapon	2.535	2.228	1096.0E6	431.7E6	427.475	340.501
2) T	3,5-DICHL...	5.490	5.123	1099.3E6	590.5E6	458.973	422.652
5) T	DICAMBA	6.464	6.136	3143.8E6	1697.1E6	457.433	424.388m
6) T	MCPD	6.630	6.229	279.4E6	139.7E6	54.825m	45.958
7) T	MCPA	6.778	6.443	345.0E6	188.6E6	47.218	44.171m
8) T	DICHLORPROP	7.154	6.759	802.7E6	460.5E6	489.298	476.665
9) T	2,4-D	7.381	7.053	995.8E6	630.8E6	547.404	538.092
10) T	Pentachlo...	7.697	7.488	9899.7E6	5344.0E6	407.213	395.694
11) T	2,4,5-TP ...	8.244	7.847	4842.8E6	3391.5E6	518.572	648.883m#
12) T	2,4,5-T	8.528	8.226	4419.6E6	2482.7E6	479.013	469.490
13) T	2,4-DB	9.081	8.735	571.2E6	321.3E6	412.030	453.816
14) T	DINOSEB	10.248	9.070	2148.2E6	1108.2E6	299.405	293.365
15) T	Picloram	10.059	10.051	4948.4E6	2981.4E6	396.450	401.307
16) T	DCPA	10.512	10.002	3712.9E6	2130.9E6	312.851	326.985

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012220\
Data File : PS008500.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2020 20:00
Operator : DD\AJ
Sample : L1004-12MS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
BU-02-101720MS

Manual Integrations
APPROVED

Ankita
1/22/2020 10:57:44 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 22 00:55:23 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010820.M
Quant Title : 8080.M
QLast Update : Wed Jan 08 11:00:59 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

