

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012422\
 Data File : PS018104.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jan 2022 10:47
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 24 20:57:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 06 10:21:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.047	6.646	1039.9E6	411.0E6	719.419	708.317
Target Compounds							
1) T	Dalapon	2.909	2.418	1536.8E6	467.0E6	626.444	610.729
2) T	3,5-DICHL...	6.287	5.862	1384.5E6	584.5E6	643.597	660.739
3) T	4-Nitroph...	6.875	6.282	592.0E6	225.3E6	732.902	734.868
5) T	DICAMBA	7.233	6.801	4045.2E6	1701.9E6	647.815	647.338
6) T	MCPD	7.402	6.896	304.4E6	136.2E6	65.207	63.341
7) T	MCPA	7.550	7.105	430.0E6	191.5E6	64.455	64.191
8) T	DICHLORPROP	7.920	7.411	1008.3E6	428.1E6	641.000	641.372
9) T	2,4-D	8.147	7.701	1211.1E6	508.9E6	665.120	691.500
10) T	Pentachlo...	8.446	8.114	13928.1E6	5474.5E6	659.288	649.997
11) T	2,4,5-TP ...	8.995	8.476	5558.9E6	2336.2E6	656.912	655.364
12) T	2,4,5-T	9.279	8.853	5855.6E6	2255.1E6	660.592	689.404
13) T	2,4-DB	9.825	9.355	1185.4E6	294.6E6	717.104	665.901
14) T	DINOSEB	10.972	9.676	4216.5E6	1610.4E6	642.076	645.436
15) T	Picloram	10.797	10.656	7880.6E6	3066.9E6	660.241	655.858
16) T	DCPA	11.240	10.602	7131.4E6	2886.9E6	660.024	653.820

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012422\
Data File : PS018104.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jan 2022 10:47
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 24 20:57:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
Quant Title : 8080.M
QLast Update : Thu Jan 06 10:21:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

