

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012422\
 Data File : PS018116.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jan 2022 21:33
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 24 21:42:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 06 10:21:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.044	6.646	1085.7E6	418.4E6	751.076	721.016
Target Compounds							
1) T	Dalapon	2.910	2.420	1561.4E6	470.9E6	636.459	615.848
2) T	3,5-DICHL...	6.285	5.862	1422.2E6	592.6E6	661.118	669.857
3) T	4-Nitroph...	6.871	6.281	610.5E6	229.4E6	755.830	748.449
5) T	DICAMBA	7.231	6.800	4119.5E6	1707.1E6	659.717	649.330
6) T	MCP	7.400	6.896	308.9E6	135.8E6	66.160	63.136
7) T	MCPA	7.548	7.105	433.3E6	190.9E6	64.947	63.989
8) T	DICHLORPROP	7.916	7.411	1026.3E6	432.2E6	652.420	647.499
9) T	2,4-D	8.144	7.700	1242.0E6	517.6E6	682.041	703.269
10) T	Pentachlo...	8.443	8.114	14249.9E6	5518.8E6	674.517	655.256
11) T	2,4,5-TP ...	8.993	8.476	5663.1E6	2370.8E6	669.229	665.075
12) T	2,4,5-T	9.275	8.852	5941.7E6	2299.0E6	670.304	702.850
13) T	2,4-DB	9.821	9.354	1228.0E6	299.9E6	742.898	677.771
14) T	DINOSEB	10.969	9.676	4302.8E6	1639.2E6	655.221	657.004
15) T	Picloram	10.793	10.654	8232.4E6	3156.5E6	689.715	675.027
16) T	DCPA	11.238	10.602	7323.6E6	2922.0E6	677.810	661.764

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012422\
Data File : PS018116.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jan 2022 21:33
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 24 21:42:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010522.M
Quant Title : 8080.M
QLast Update : Thu Jan 06 10:21:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

