

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012423\
 Data File : PS021862.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jan 2023 11:17
 Operator : AR\AJ
 Sample : PB150425BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB150425BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 24 20:35:13 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010923.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 10 10:34:52 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.378	6.810	1382.4E6	871.7E6	496.616	485.173
Target Compounds							
1) T	Dalapon	2.483	2.557	1240.9E6	1091.5E6	438.651	424.004
2) T	3,5-DICHL...	5.749	5.992	1807.9E6	1248.9E6	467.359	453.190
3) T	4-Nitroph...	6.222	6.430	744.3E6	516.4E6	510.601	471.106
5) T	DICAMBA	6.517	6.972	5774.1E6	3905.1E6	472.238	459.279
6) T	MCPD	6.670	7.067	424.7E6	288.5E6	44.321	43.505
7) T	MCPA	6.795	7.280	559.4E6	395.0E6	43.310	43.553
8) T	DICHLORPROP	7.115	7.598	1501.1E6	984.2E6	465.431	461.318
9) T	2,4-D	7.322	7.892	1789.8E6	1161.5E6	460.840	459.721
10) T	Pentachlo...	7.544	8.315	21762.0E6	14061.1E6	522.443	489.957
11) T	2,4,5-TP ...	8.074	8.678	9013.3E6	5642.6E6	489.162	474.552
12) T	2,4,5-T	8.335	9.054	9400.6E6	5385.3E6	487.691	482.123
13) T	2,4-DB	8.847	9.562	1468.8E6	721.6E6	464.020	481.652
14) T	DINOSEB	9.894	9.887	7891.6E6	3844.0E6	480.799	462.531
15) T	Picloram	9.744	10.865	13146.5E6	7464.2E6	463.436	421.507
16) T	DCPA	10.161	10.815	12033.2E6	6754.2E6	487.101	484.530

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012423\
Data File : PS021862.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jan 2023 11:17
Operator : AR\AJ
Sample : PB150425BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB150425BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 24 20:35:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010923.M
Quant Title : 8080.M
QLast Update : Tue Jan 10 10:34:52 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

