

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012720\
 Data File : PS008555.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jan 2020 14:55
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 04:46:33 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010820.M
 Quant Title : 8080.M
 QLast Update : Wed Jan 08 11:00:59 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.276	5.967	1344.0E6	636.7E6	761.148	671.977
Target Compounds							
1) T	Dalapon	2.537	2.225	1698.7E6	770.3E6	662.575	607.610
2) T	3,5-DICHL...	5.494	5.123	1689.5E6	878.9E6	705.372	629.033
3) T	4-Nitroph...	6.110	5.574	631.3E6	355.9E6	648.143	599.351
5) T	DICAMBA	6.469	6.137	5015.7E6	2574.9E6	729.802	643.891
6) T	MCPD	6.637	6.232	371.3E6	191.6E6	72.854	63.000
7) T	MCPA	6.786	6.447	526.3E6	255.5E6	72.021	59.842
8) T	DICHLORPROP	7.159	6.761	1183.4E6	616.9E6	721.367	638.481
9) T	2,4-D	7.386	7.054	1320.2E6	741.4E6	725.736	632.489
10) T	Pentachlo...	7.702	7.490	17661.9E6	8817.0E6	726.505	652.850
11) T	2,4,5-TP ...	8.249	7.849	6841.6E6	3407.9E6	732.603	652.033
12) T	2,4,5-T	8.533	8.228	6570.5E6	3411.4E6	712.128	645.098
13) T	2,4-DB	9.086	8.736	933.6E6	447.7E6	673.502	632.436
14) T	DINOSEB	10.253	9.072	5017.5E6	2428.7E6	699.313	642.910
15) T	Picloram	10.065	10.053	8573.6E6	4874.1E6	686.897	656.062
16) T	DCPA	10.518	10.004	8532.1E6	4382.9E6	718.925	672.574

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012720\
Data File : PS008555.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jan 2020 14:55
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 28 04:46:33 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010820.M
Quant Title : 8080.M
QLast Update : Wed Jan 08 11:00:59 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

