

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012721\
 Data File : PS013872.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jan 2021 18:35
 Operator : DD\AJ
 Sample : HSTDICC750
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 00:53:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012721.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 28 00:52:07 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.203	6.906	812.9E6	290.7E6	750.000	750.000
Target Compounds							
1) T	Dalapon	3.026	2.621	863.7E6	352.7E6	682.500	682.500
2) T	3,5-DICHL...	6.420	6.062	1046.5E6	424.8E6	697.500	697.500
3) T	4-Nitroph...	7.037	6.520	377.0E6	174.9E6	682.500	682.500
5) T	DICAMBA	7.392	7.072	3079.0E6	1181.3E6	705.000	705.000
6) T	MCPD	7.559	7.167	229.9E6	95708208	70.500	70.500
7) T	MCPA	7.709	7.381	322.4E6	135.8E6	69.750	69.750
8) T	DICHLORPROP	8.083	7.697	770.9E6	301.3E6	705.000	705.000
9) T	2,4-D	8.315	7.995	901.6E6	380.2E6	705.000	705.000
10) T	Pentachlo...	8.621	8.418	11488.8E6	4022.4E6	712.500	712.500
11) T	2,4,5-TP ...	9.170	8.780	4550.0E6	1665.6E6	712.500	712.500
12) T	2,4,5-T	9.457	9.161	4684.6E6	1633.7E6	712.500	712.500
13) T	2,4-DB	10.011	9.668	808.5E6	219.8E6	712.500	712.500
14) T	DINOSEB	11.162	9.995	3621.7E6	1179.7E6	705.000	705.000
15) T	Picloram	10.997	10.990	6338.5E6	2522.8E6	712.500	712.500
16) T	DCPA	11.427	10.921	5912.6E6	2194.2E6	720.000	720.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012721\
Data File : PS013872.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jan 2021 18:35
Operator : DD\AJ
Sample : HSTDICC750
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 28 00:53:11 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012721.M
Quant Title : 8080.M
QLast Update : Thu Jan 28 00:52:07 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

