

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\
 Method File : PS012819.M
 Title : 8080.M
 Last Update : Tue Jan 29 06:11:45 2019
 Response Via : Initial Calibration

Calibration Files

750 =PS002946.D 1000 =PS002947.D 1500 =PS002948.D
 500 =PS002945.D 200 =PS002944.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.440	5.454	5.389	5.487	5.813	5.516	E5	3.07
2) T	3,5-DICHLOROBENZOI...	5.710	5.726	5.498	5.610	6.247	5.758	E5	5.00
3) T	4-Nitrophenol	1.984	2.037	1.886	1.716	2.054	1.935	E5	7.18
4) S	2,4-DCAA	3.794	3.826	3.671	3.633	4.059	3.796	E5	4.41
5) T	DICAMBA	1.606	1.617	1.575	1.585	1.630	1.603	E6	1.41
6) T	MCPP	1.191	1.233	1.259	1.155	1.152	1.198	E6	3.95
7) T	MCPA	1.767	1.814	1.837	1.673	1.760	1.770	E6	3.57
8) T	DICHLORPROP	3.793	3.807	3.698	3.818	4.060	3.835	E5	3.50
9) T	2,4-D	4.656	4.753	4.618	4.520	4.506	4.611	E5	2.21
10) T	Pentachlorophenol	6.454	6.315	5.878	6.529	6.889	6.413	E6	5.72
11) T	2,4,5-TP (SILVEX)	2.523	2.516	2.414	2.527	2.613	2.518	E6	2.80
12) T	2,4,5-T	2.440	2.445	2.361	2.455	2.496	2.439	E6	2.02
13) T	2,4-DB	2.677	2.738	2.725	2.457	2.474	2.614	E5	5.27
14) T	DINOSEB	1.862	1.852	1.785	1.897	2.010	1.881	E6	4.40
15) T	Picloram	3.201	3.276	3.244	3.126	2.972	3.164	E6	3.83
16) T	DCPA	3.011	2.970	2.811	3.035	3.208	3.007	E6	4.73

Signal #2 Calibration Files

750 =PS002946.D 1000 =PS002947.D 1500 =PS002948.D
 500 =PS002945.D 200 =PS002944.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.336	5.377	5.143	5.285	5.654	5.359	E5	3.49
2) T	3,5-DICHLOROBENZOI...	5.554	5.573	5.463	5.576	6.013	5.636	E5	3.83
3) T	4-Nitrophenol	2.027	2.011	2.016	2.010	2.387	2.090	E5	7.95
4) S	2,4-DCAA	4.247	4.272	4.095	4.205	4.558	4.275	E5	4.02
5) T	DICAMBA	1.587	1.595	1.551	1.577	1.623	1.587	E6	1.66
6) T	MCPP	1.108	1.141	1.152	1.058	1.055	1.103	E6	4.11
7) T	MCPA	1.599	1.633	1.669	1.512	1.452	1.573	E6	5.66
8) T	DICHLORPROP	3.755	3.770	3.716	3.795	4.082	3.823	E5	3.85
9) T	2,4-D	4.981	5.030	4.926	4.939	5.192	5.014	E5	2.15
10) T	Pentachlorophenol	6.190	6.043	5.625	6.343	6.702	6.181	E6	6.40
11) T	2,4,5-TP (SILVEX)	2.480	2.475	2.355	2.488	2.604	2.480	E6	3.55
12) T	2,4,5-T	1.928	1.898	1.765	1.920	2.057	1.913	E6	5.43
13) T	2,4-DB	2.503	2.626	2.625	2.518	2.428	2.540	E5	3.35
14) T	DINOSEB	1.751	1.746	1.688	1.746	1.834	1.753	E6	2.99
15) T	Picloram	3.036	3.103	3.079	2.997	2.816	3.006	E6	3.78
16) T	DCPA	3.076	3.041	2.884	3.158	3.329	3.098	E6	5.27

(#) = Out of Range