

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012819\
 Data File : PS002944.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jan 2019 16:56
 Operator : EI\SJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Manual Integrations
 APPROVED

Sohil
 1/29/2019 11:33:26 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 05:46:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012819.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 29 05:45:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.317	10.815	81172398	91159385	213.974	214.626
Target Compounds							
1) T	Dalapon	2.876	2.344	105.8E6	102.9E6	194.496	192.857m
2) T	3,5-DICHL...	9.745	8.975	116.2E6	111.8E6	203.475	201.378
3) T	4-Nitroph...	10.915	9.969	37386445	43445111	188.418	214.290m
5) T	DICAMBA	11.621	11.094	306.5E6	305.1E6	190.882	192.221
6) T	MCPD	11.960	11.320	21664775	19842635	18.187m	17.912m
7) T	MCPA	12.209	11.695	32734947	27016349	18.531m	16.894
8) T	DICHLORPROP	12.813	12.220	76329179	76734318	201.215	204.339
9) T	2,4-D	13.177	12.694	84708465	97616148	181.922	195.978
10) T	Pentachlo...	13.540	13.227	1308.9E6	1273.3E6	202.812	205.725
11) T	2,4,5-TP ...	14.417	13.852	496.4E6	494.7E6	196.714	199.447
12) T	2,4,5-T	14.817	14.417	474.3E6	390.9E6	194.369	202.775
13) T	2,4-DB	15.565	15.070	47009923	46140579	175.618	184.348
14) T	DINOSEB	17.032	15.492	377.9E6	344.9E6	203.030	196.938
15) T	Picloram	16.808	16.723	564.6E6	535.1E6	176.391	176.230
16) T	DCPA	17.377	16.666	616.0E6	639.1E6	204.569	207.769m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012819\
Data File : PS002944.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jan 2019 16:56
Operator : EI\SJ
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDICC200

Manual Integrations
APPROVED

Sohil
1/29/2019 11:33:26 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 05:46:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012819.M
Quant Title : 8080.M
QLast Update : Tue Jan 29 05:45:06 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

