

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012819\
 Data File : PS002945.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jan 2019 17:28
 Operator : EI\SJ
 Sample : HSTDICC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC500

Manual Integrations
 APPROVED

Sohil
 1/29/2019 11:33:30 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 05:46:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012819.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 29 05:45:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.280	10.779	181.6E6	210.2E6	478.811m	494.956m
Target Compounds							
1) T	Dalapon	2.842	2.311	249.7E6	240.5E6	458.973m	450.673m
2) T	3,5-DICHL...	9.708	8.939	260.9E6	259.3E6	456.858m	466.842m
3) T	4-Nitroph...	10.869	9.929	78083997	91451929	393.522m	451.081m
5) T	DICAMBA	11.586	11.059	745.1E6	741.4E6	463.996m	467.147m
6) T	MCPD	11.930	11.290	54276074	49721321	45.563m	44.883m
7) T	MCPA	12.179	11.664	77775902	70300023	44.028m	43.960m
8) T	DICHLORPROP	12.777	12.185	179.4E6	178.4E6	472.997m	474.959
9) T	2,4-D	13.138	12.657	212.4E6	232.1E6	456.241m	466.064m
10) T	Pentachlo...	13.505	13.192	3101.3E6	3013.1E6	480.530m	486.796m
11) T	2,4,5-TP ...	14.383	13.817	1200.2E6	1182.0E6	475.617	476.531m
12) T	2,4,5-T	14.781	14.380	1166.1E6	912.0E6	477.857	473.119m
13) T	2,4-DB	15.527	15.033	116.7E6	119.6E6	435.965m	477.831m
14) T	DINOSEB	16.998	15.457	891.7E6	820.6E6	478.986	468.625m
15) T	Picloram	16.769	16.687	1484.8E6	1423.6E6	463.875	468.837
16) T	DCPA	17.343	16.632	1457.0E6	1516.0E6	483.830m	492.874m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012819\
Data File : PS002945.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jan 2019 17:28
Operator : EI\SJ
Sample : HSTDICC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDICC500

Manual Integrations
APPROVED

Sohil
1/29/2019 11:33:30 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 05:46:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012819.M
Quant Title : 8080.M
QLast Update : Tue Jan 29 05:45:06 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

