

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012820\
 Data File : PS008588.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jan 2020 19:44
 Operator : DD\AJ
 Sample : L1008-21MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 CL-02-012420MSD

Manual Integrations
 APPROVED

Sohil
 1/30/2020 3:29:51 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 04:40:24 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010820.M
 Quant Title : 8080.M
 QLast Update : Wed Jan 08 11:00:59 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.263	5.941	539.1E6	267.0E6	305.290m	281.846m
Target Compounds							
1) T	Dalapon	2.528	2.205	1041.8E6	359.3E6	406.346	283.385 #
2) T	3,5-DICHL...	5.481	5.099	1098.7E6	556.2E6	458.727m	398.116m
5) T	DICAMBA	6.456	6.205	3522.1E6	145.2E6	512.471	36.303m#
6) T	MCP	6.622	6.205	332.7E6	150.4E6	65.279m	49.466m
7) T	MCPA	6.770	6.420	407.2E6	230.4E6	55.718m	53.965m
8) T	DICHLORPROP	7.146	6.735	942.8E6	526.3E6	574.686	544.781
9) T	2,4-D	7.374	7.029	1151.9E6	641.1E6	633.197	546.871m
10) T	Pentachlo...	7.688	7.464	8315.9E6	4206.6E6	342.065m	311.480m
11) T	2,4,5-TP ...	8.236	7.822	5642.9E6	3231.4E6	604.248	618.256m
12) T	2,4,5-T	8.520	8.202	5212.0E6	2735.5E6	564.895	517.291m
13) T	2,4-DB	9.073	8.711	687.5E6	355.1E6	495.973	501.679
14) T	DINOSEB	10.239	9.045	1751.2E6	811.2E6	244.071m	214.723m
15) T	Picloram	10.050	10.026	4676.8E6	2552.9E6	374.695m	343.634m
16) T	DCPA	10.504	9.977	4186.9E6	2225.8E6	352.797m	341.560m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012820\
 Data File : PS008588.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jan 2020 19:44
 Operator : DD\AJ
 Sample : L1008-21MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleID :
 CL-02-012420MSD

Manual Integrations
 APPROVED

Sohil
 1/30/2020 3:29:51 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 04:40:24 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010820.M
 Quant Title : 8080.M
 QLast Update : Wed Jan 08 11:00:59 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

