

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012820\
 Data File : PS008604.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jan 2020 01:35
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 1/30/2020 3:30:07 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 04:42:13 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010820.M
 Quant Title : 8080.M
 QLast Update : Wed Jan 08 11:00:59 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.261	5.955	1391.8E6	667.1E6	788.216	704.071m
Target Compounds							
1) T	Dalapon	2.527	2.220	1723.7E6	720.4E6	672.296	568.216
2) T	3,5-DICHL...	5.480	5.114	1781.2E6	923.5E6	743.673m	660.988m
3) T	4-Nitroph...	6.096	5.563	681.4E6	374.1E6	699.613	630.060m
5) T	DICAMBA	6.455	6.126	5299.0E6	2686.7E6	771.025	671.838m
6) T	MCPD	6.622	6.221	409.9E6	198.8E6	80.430m	65.388
7) T	MCPA	6.770	6.436	562.4E6	285.5E6	76.965	66.869m
8) T	DICHLORPROP	7.144	6.749	1256.9E6	647.4E6	766.121	670.079m
9) T	2,4-D	7.371	7.043	1348.0E6	739.0E6	741.005	630.443m
10) T	Pentachlo...	7.687	7.478	18525.6E6	9108.2E6	762.033m	674.419m
11) T	2,4,5-TP ...	8.234	7.837	7229.6E6	3550.2E6	774.151	679.247m
12) T	2,4,5-T	8.517	8.217	7016.6E6	3581.5E6	760.481	677.279m
13) T	2,4-DB	9.070	8.725	974.7E6	472.4E6	703.092m	667.277
14) T	DINOSEB	10.237	9.061	5269.8E6	2503.7E6	734.483m	662.758
15) T	Picloram	10.049	10.041	7624.4E6	4189.2E6	610.847m	563.884
16) T	DCPA	10.501	9.992	8932.0E6	4571.0E6	752.623m	701.425m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012820\
Data File : PS008604.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jan 2020 01:35
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
1/30/2020 3:30:07 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 04:42:13 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS010820.M
Quant Title : 8080.M
QLast Update : Wed Jan 08 11:00:59 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

