

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012821\
 Data File : PS013880.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jan 2021 14:56
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 15:55:52 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012721.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 28 01:00:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.203	6.906	818.1E6	292.6E6	760.393	758.571
Target Compounds							
1) T	Dalapon	3.027	2.621	846.2E6	347.4E6	678.042	675.656
2) T	3,5-DICHL...	6.420	6.062	1043.9E6	430.0E6	702.457	708.262
3) T	4-Nitroph...	7.038	6.520	369.9E6	171.3E6	680.805	681.378
5) T	DICAMBA	7.392	7.072	3122.8E6	1202.5E6	726.272	729.677
6) T	MCPD	7.559	7.166	234.5E6	98008016	74.317	72.972
7) T	MCPA	7.708	7.381	328.4E6	138.6E6	72.561	71.515
8) T	DICHLORPROP	8.083	7.696	778.0E6	306.7E6	715.737	720.808
9) T	2,4-D	8.315	7.995	901.4E6	382.3E6	734.068	717.180
10) T	Pentachlo...	8.621	8.418	11481.5E6	4052.0E6	725.008	727.372
11) T	2,4,5-TP ...	9.170	8.779	4542.9E6	1678.7E6	723.519	731.298
12) T	2,4,5-T	9.456	9.160	4646.1E6	1632.6E6	721.323	733.171
13) T	2,4-DB	10.010	9.666	784.8E6	218.4E6	710.968	730.459
14) T	DINOSEB	11.161	9.994	3519.0E6	1161.3E6	694.352	710.865
15) T	Picloram	10.994	10.988	6239.4E6	2457.7E6	734.808	720.167
16) T	DCPA	11.427	10.920	5805.7E6	2192.8E6	718.189	731.951

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012821\
Data File : PS013880.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jan 2021 14:56
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 28 15:55:52 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012721.M
Quant Title : 8080.M
QLast Update : Thu Jan 28 01:00:30 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

