

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012925\
 Data File : PS028986.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jan 2025 19:27
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 30 00:43:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 14 12:25:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.196	7.674	2119.5E6	804.0E6	761.297	720.514
Target Compounds							
1) T	Dalapon	2.615	2.665	2209.2E6	1311.4E6	740.910	642.773
2) T	3,5-DICHL...	6.372	6.640	2881.1E6	1112.4E6	720.846	673.092
3) T	4-Nitroph...	6.993	7.204	1251.6E6	619.8E6	706.263	696.585
5) T	DICAMBA	7.381	7.871	8797.6E6	4040.5E6	741.696	725.527
6) T	MCPD	7.563	7.976	537.9E6	198.2E6	78.907	65.891
7) T	MCPA	7.712	8.217	729.2E6	271.3E6	74.021	63.874
8) T	DICHLORPROP	8.084	8.582	2262.5E6	985.8E6	714.051	701.327
9) T	2,4-D	8.314	8.909	2433.2E6	1039.4E6	720.000	693.151
10) T	Pentachlo...	8.611	9.432	35907.0E6	17336.1E6	744.388	748.344
11) T	2,4,5-TP ...	9.186	9.809	14025.3E6	7020.6E6	733.053	745.335
12) T	2,4,5-T	9.477	10.225	14119.5E6	6616.9E6	735.508	734.476
13) T	2,4-DB	10.047	10.790	2590.4E6	697.8E6	730.247	700.813
14) T	DINOSEB	11.251	11.168	11621.0E6	4533.4E6	702.292	706.433
15) T	Picloram	11.061	12.250	22663.4E6	9889.7E6	718.283	736.956
16) T	DCPA	11.547	12.206	21190.1E6	8788.5E6	738.820	774.160

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS012925\
Data File : PS028986.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jan 2025 19:27
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 30 00:43:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

