

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013024\
 Data File : PS025682.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jan 2024 15:44
 Operator : AR\AJ
 Sample : P1243-02MSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MR-308-0-2MSD

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 01/31/2024
 Supervised By : Ankita Jodhani 01/31/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 31 02:04:30 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011124.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 11 16:45:24 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.349	6.756	863.9E6	378.7E6	580.029	493.248m
Target Compounds							
1) T	Dalapon	2.476	2.531	815.2E6	423.5E6	352.260	360.516
2) T	3,5-DICHL...	5.725	5.950	1082.7E6	521.6E6	473.126	447.974
5) T	DICAMBA	6.488	6.915	2760.3E6	1417.9E6	432.562	426.061
6) T	MCP	6.637	7.011	185.2E6	113.0E6	46.253	48.705
7) T	MCPA	6.758	7.215	263.3E6	198.9E6	46.519m	57.400m
8) T	DICHLORPROP	7.073	7.527	781.5E6	413.6E6	474.248	478.458
9) T	2,4-D	7.271	7.809	976.6E6	544.5E6	530.347m	599.362
10) T	Pentachlo...	7.498	8.229	11870.9E6	5908.7E6	496.052	480.381m
11) T	2,4,5-TP ...	8.019	8.589	4745.1E6	2602.0E6	510.566	551.181
12) T	2,4,5-T	8.275	8.958	4759.0E6	2281.7E6	487.416	522.052m
13) T	2,4-DB	8.781	9.460	691.2E6	292.1E6	642.113	559.597m
14) T	DINOSEB	9.828	9.788	3070.0E6	1534.1E6	447.688	487.994
15) T	Picloram	9.663	10.739	4770.2E6	2238.3E6	461.060	420.988
16) T	DCPA	10.093	10.707	5873.4E6	2466.1E6	484.241m	424.140

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013024\
Data File : PS025682.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2024 15:44
Operator : AR\AJ
Sample : P1243-02MSD
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
MR-308-0-2MSD

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 01/31/2024
Supervised By : Ankita Jodhani 01/31/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 02:04:30 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011124.M
Quant Title : 8080.M
QLast Update : Thu Jan 11 16:45:24 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

