

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013024\
 Data File : PS025686.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jan 2024 17:44
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 01/31/2024
 Supervised By :Ankita Jodhani 01/31/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 31 02:06:55 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011124.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 11 16:45:24 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.348	6.756	1147.7E6	587.1E6	770.555	764.644
Target Compounds							
1) T	Dalapon	2.476	2.531	1562.8E6	763.1E6	675.359	649.534
2) T	3,5-DICHL...	5.725	5.950	1640.3E6	832.4E6	716.798	715.003
3) T	4-Nitroph...	6.188	6.376	658.0E6	334.6E6	728.203	702.834
5) T	DICAMBA	6.488	6.915	4719.2E6	2451.3E6	739.531	736.583
6) T	MCPD	6.638	7.012	307.6E6	172.1E6	76.827	74.188
7) T	MCPA	6.760	7.217	429.9E6	249.9E6	75.965	72.128
8) T	DICHLORPROP	7.073	7.527	1201.5E6	618.5E6	729.092	715.503
9) T	2,4-D	7.272	7.809	1298.7E6	668.8E6	705.301	736.207
10) T	Pentachlo...	7.498	8.230	17960.6E6	9131.5E6	750.527	742.403
11) T	2,4,5-TP ...	8.020	8.589	6983.4E6	3558.9E6	751.403	753.898
12) T	2,4,5-T	8.276	8.958	6970.5E6	3340.5E6	713.926m	764.290
13) T	2,4-DB	8.782	9.461	850.0E6	397.1E6	789.623	760.603
14) T	DINOSEB	9.828	9.787	5180.5E6	2343.1E6	755.456	745.297
15) T	Picloram	9.667	10.741	8227.5E6	3882.7E6	795.232	730.265
16) T	DCPA	10.093	10.706	9249.8E6	4217.6E6	762.617	725.377

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013024\
Data File : PS025686.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2024 17:44
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 01/31/2024
Supervised By : Ankita Jodhani 01/31/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 02:06:55 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011124.M
Quant Title : 8080.M
QLast Update : Thu Jan 11 16:45:24 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

