

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
 Data File : PS028989.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jan 2025 13:20
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 31 05:20:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 14 12:25:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.192	7.670	2165.4E6	822.3E6	777.786	736.959
Target Compounds							
1) T	Dalapon	2.613	2.665	2223.3E6	1302.6E6	745.627	638.476
2) T	3,5-DICHL...	6.370	6.637	2940.8E6	1137.4E6	735.774	688.221
3) T	4-Nitroph...	6.990	7.200	1301.1E6	634.4E6	734.214	713.021
5) T	DICAMBA	7.377	7.867	9036.6E6	4147.8E6	761.847	744.805
6) T	MCP	7.560	7.972	530.1E6	205.7E6	77.774	68.381
7) T	MCPA	7.708	8.213	755.5E6	281.9E6	76.696	66.376
8) T	DICHLORPROP	8.081	8.578	2319.8E6	1009.7E6	732.116	718.335
9) T	2,4-D	8.310	8.905	2484.7E6	1066.3E6	735.236	711.082
10) T	Pentachlo...	8.606	9.427	36863.8E6	17810.8E6	764.221	768.832
11) T	2,4,5-TP ...	9.181	9.803	14362.5E6	7220.6E6	750.677	766.568
12) T	2,4,5-T	9.472	10.219	14459.2E6	6819.7E6	753.206	756.985
13) T	2,4-DB	10.042	10.784	2641.6E6	719.5E6	744.655	722.531
14) T	DINOSEB	11.245	11.162	12050.0E6	4752.1E6	728.215	740.508
15) T	Picloram	11.055	12.243	23546.5E6	10381.1E6	746.273	773.575
16) T	DCPA	11.540	12.199	21648.2E6	9068.6E6	754.790	798.830

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS013025\
Data File : PS028989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 13:20
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 05:20:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011425.M
Quant Title : 8080.M
QLast Update : Tue Jan 14 12:25:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

