

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020223\
 Data File : PS021985.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Feb 2023 18:53
 Operator : AR\AJ
 Sample : HSTDICC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 03 01:48:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020223.M
 Quant Title : 8080.M
 QLast Update : Fri Feb 03 01:41:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.304	6.732	1146.1E6	989.3E6	513.652	496.271
Target Compounds							
1) T	Dalapon	2.412	2.486	1392.7E6	1403.3E6	447.139	449.565
2) T	3,5-DICHL...	5.684	5.929	1619.6E6	1366.7E6	477.223	463.284
3) T	4-Nitroph...	6.152	6.358	605.0E6	600.2E6	443.705	449.885
5) T	DICAMBA	6.439	6.889	4625.1E6	4514.6E6	484.555	482.905
6) T	MCP	6.588	6.986	331.6E6	343.7E6	44.880	45.083
7) T	MCPA	6.709	7.195	460.8E6	459.9E6	45.469	45.520
8) T	DICHLORPROP	7.024	7.511	1165.6E6	1116.9E6	481.397	473.444
9) T	2,4-D	7.230	7.803	1437.2E6	1324.4E6	488.782	475.854
10) T	Pentachlo...	7.444	8.218	16738.1E6	16002.3E6	530.790	526.898
11) T	2,4,5-TP ...	7.973	8.584	6471.2E6	6408.0E6	497.218	497.514
12) T	2,4,5-T	8.235	8.960	7048.0E6	5838.7E6	506.898	485.361
13) T	2,4-DB	8.742	9.465	1055.2E6	795.1E6	449.905	469.249
14) T	DINOSEB	9.781	9.787	4914.0E6	4441.6E6	492.305	480.489
15) T	Picloram	9.625	10.759	8820.0E6	8742.3E6	466.408	478.924
16) T	DCPA	10.050	10.712	8388.0E6	7599.2E6	508.341	509.495

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020223\
Data File : PS021985.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Feb 2023 18:53
Operator : AR\AJ
Sample : HSTDICC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 03 01:48:27 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020223.M
Quant Title : 8080.M
QLast Update : Fri Feb 03 01:41:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

