

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020223\
 Data File : PS021986.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Feb 2023 19:17
 Operator : AR\AJ
 Sample : HSTDICC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 03 01:42:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020223.M
 Quant Title : 8080.M
 QLast Update : Fri Feb 03 01:41:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.302	6.730	1694.2E6	1508.0E6	750.000	750.000
Target Compounds							
1) T	Dalapon	2.412	2.484	2144.4E6	2144.0E6	682.500	682.500
2) T	3,5-DICHL...	5.682	5.928	2414.9E6	2093.5E6	697.500	697.500
3) T	4-Nitroph...	6.145	6.354	935.4E6	917.0E6	682.500	682.500
5) T	DICAMBA	6.438	6.888	6885.6E6	6733.0E6	705.000	705.000
6) T	MCP	6.590	6.988	522.6E6	537.2E6	70.500	70.500
7) T	MCPA	6.713	7.198	710.8E6	707.8E6	69.750	69.750
8) T	DICHLORPROP	7.022	7.509	1728.8E6	1680.4E6	705.000	705.000
9) T	2,4-D	7.225	7.799	2132.4E6	1985.3E6	705.000	705.000
10) T	Pentachlo...	7.443	8.218	23850.4E6	22689.8E6	712.500	712.500
11) T	2,4,5-TP ...	7.971	8.583	9523.8E6	9420.8E6	712.500	712.500
12) T	2,4,5-T	8.231	8.957	10164.0E6	8786.6E6	712.500	712.500
13) T	2,4-DB	8.737	9.463	1650.0E6	1215.6E6	712.500	712.500
14) T	DINOSEB	9.781	9.786	7181.3E6	6635.4E6	705.000	705.000
15) T	Picloram	9.620	10.756	13583.7E6	13159.2E6	712.500	712.500
16) T	DCPA	10.049	10.711	12192.3E6	11050.1E6	720.000	720.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020223\
Data File : PS021986.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Feb 2023 19:17
Operator : AR\AJ
Sample : HSTDICC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 03 01:42:06 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020223.M
Quant Title : 8080.M
QLast Update : Fri Feb 03 01:41:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

