

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020323\
 Data File : PS022022.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Feb 2023 12:43
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 03 23:36:04 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020223.M
 Quant Title : 8080.M
 QLast Update : Fri Feb 03 10:03:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.302	6.727	1646.5E6	1473.8E6	722.360	728.098
Target Compounds							
1) T	Dalapon	2.413	2.485	2040.2E6	2033.5E6	657.083	653.505
2) T	3,5-DICHL...	5.682	5.925	2333.3E6	2058.3E6	677.574	688.807
3) T	4-Nitroph...	6.143	6.351	899.8E6	891.6E6	668.534	667.672
5) T	DICAMBA	6.439	6.886	6723.0E6	6614.1E6	695.048	700.080
6) T	MCP	6.592	6.986	508.1E6	526.3E6	70.560	70.521
7) T	MCPA	6.714	7.196	693.0E6	693.7E6	69.202	69.061
8) T	DICHLORPROP	7.022	7.507	1688.9E6	1654.1E6	683.148	691.353
9) T	2,4-D	7.223	7.796	1944.9E6	1950.1E6	646.918	691.598
10) T	Pentachlo...	7.444	8.216	23444.8E6	22365.1E6	710.031	703.962
11) T	2,4,5-TP ...	7.972	8.580	9324.5E6	9256.0E6	699.866	702.781
12) T	2,4,5-T	8.231	8.954	9789.2E6	8708.0E6	682.101	720.574
13) T	2,4-DB	8.736	9.459	1630.7E6	1193.1E6	741.678	706.559
14) T	DINOSEB	9.782	9.783	6967.1E6	6485.5E6	677.454	693.833
15) T	Picloram	9.618	10.752	13168.2E6	12762.7E6	707.078	703.221
16) T	DCPA	10.050	10.708	11986.1E6	10941.4E6	705.433	710.941

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020323\
Data File : PS022022.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Feb 2023 12:43
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 03 23:36:04 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020223.M
Quant Title : 8080.M
QLast Update : Fri Feb 03 10:03:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

