

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020419\
 Data File : PS003015.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Feb 2019 00:39
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 2/5/2019 11:25:56 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 05 06:08:40 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012819.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 29 06:11:45 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.298	10.780	301.0E6	323.4E6	792.810	756.449m
Target Compounds							
1) T	Dalapon	2.859	2.314	397.6E6	369.0E6	720.769m	688.546
2) T	3,5-DICHL...	9.725	8.940	417.6E6	397.1E6	725.124m	704.652m
3) T	4-Nitroph...	10.884	9.929	157.1E6	155.0E6	811.974m	741.777m
5) T	DICAMBA	11.604	11.061	1156.6E6	1129.8E6	721.601m	712.076m
6) T	MCP	11.949	11.295	90014804	82254792	75.130m	74.577m
7) T	MCPA	12.199	11.670	135.6E6	120.9E6	76.616m	76.834m
8) T	DICHLORPROP	12.794	12.186	279.9E6	275.3E6	729.731m	719.958
9) T	2,4-D	13.152	12.657	337.7E6	363.1E6	732.496m	724.223m
10) T	Pentachlo...	13.523	13.193	4680.6E6	4439.7E6	729.858m	718.330m
11) T	2,4,5-TP ...	14.399	13.818	1838.7E6	1773.9E6	730.120	715.196m
12) T	2,4,5-T	14.797	14.381	1782.0E6	1302.8E6	730.505	680.859m
13) T	2,4-DB	15.543	15.033	202.3E6	191.3E6	773.663m	753.219m
14) T	DINOSEB	17.015	15.458	1344.6E6	1241.7E6	714.697	708.330m
15) T	Picloram	16.784	16.688	2349.6E6	2243.9E6	742.689	746.357
16) T	DCPA	17.359	16.633	2192.2E6	2244.8E6	728.971m	724.685m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020419\
Data File : PS003015.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Feb 2019 00:39
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
2/5/2019 11:25:56 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 06:08:40 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012819.M
Quant Title : 8080.M
QLast Update : Tue Jan 29 06:11:45 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

