

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020421\
 Data File : PS013939.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Feb 2021 17:44
 Operator : DD\AJ
 Sample : M1293-02MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TEST-PIT-4MSD

Manual Integrations
 APPROVED

Ankita
 2/8/2021 8:43:19 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 05 00:27:12 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012721.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 28 01:00:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.205	6.904	433.3E6	192.5E6	402.724	499.146
Target Compounds							
1) T	Dalapon	3.025	2.621	566.6E6	304.1E6	454.007	591.511 #
2) T	3,5-DICHL...	6.421	6.060	701.7E6	292.4E6	472.226	481.640
5) T	DICAMBA	7.394	7.070	1968.4E6	877.1E6	457.798	532.261
6) T	MCP	7.560	7.163	160.2E6	73309848	50.769	54.583
7) T	MCPA	7.708	7.377	275.4E6	125.4E6	60.856	64.663
8) T	DICHLORPROP	8.085	7.695	515.2E6	235.3E6	474.023	553.133
9) T	2,4-D	8.316	7.991	619.5E6	300.3E6	504.531	563.455
10) T	Pentachlo...	8.622	8.416	6437.9E6	2623.0E6	406.524	470.866
11) T	2,4,5-TP ...	9.170	8.775	3215.8E6	1674.4E6	512.163	729.429m#
12) T	2,4,5-T	9.457	9.157	3143.2E6	1277.0E6	487.996	573.482
13) T	2,4-DB	10.009	9.663	399.9E6	163.2E6	362.295	545.786 #
14) T	DINOSEB	11.164	9.992	1619.3E6	598.4E6	319.521	366.292
15) T	Picloram	10.996	10.986	3161.3E6	1462.7E6	372.310	428.620
16) T	DCPA	11.429	10.918	3713.6E6	1627.2E6	459.378	543.141

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020421\
Data File : PS013939.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Feb 2021 17:44
Operator : DD\AJ
Sample : M1293-02MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
TEST-PIT-4MSD

Manual Integrations
APPROVED

Ankita
2/8/2021 8:43:19 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 00:27:12 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012721.M
Quant Title : 8080.M
QLast Update : Thu Jan 28 01:00:30 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

