

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020519\
 Data File : PS003029.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Feb 2019 16:32
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleID :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 2/6/2019 4:17:58 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 06 00:13:53 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012819.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 29 06:11:45 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.321	10.820	297.0E6	329.8E6	782.222	771.479
Target Compounds							
1) T	Dalapon	2.883	2.354	393.3E6	395.6E6	712.908	738.210
2) T	3,5-DICHL...	9.748	8.980	416.8E6	401.4E6	723.858	712.255
3) T	4-Nitroph...	10.907	9.968	150.8E6	152.0E6	779.233	727.318m
5) T	DICAMBA	11.627	11.101	1159.7E6	1142.9E6	723.478	720.334
6) T	MCPD	11.971	11.334	90149991	83456546	75.243m	75.667m
7) T	MCPA	12.222	11.710	134.1E6	121.6E6	75.752	77.308m
8) T	DICHLORPROP	12.818	12.226	279.0E6	275.4E6	727.413	720.417
9) T	2,4-D	13.177	12.697	346.8E6	366.8E6	752.115	731.564
10) T	Pentachlo...	13.546	13.233	4725.7E6	4496.1E6	736.891	727.460
11) T	2,4,5-TP ...	14.423	13.857	1853.4E6	1805.4E6	735.954	727.867m
12) T	2,4,5-T	14.820	14.420	1789.9E6	1384.8E6	733.741	723.719m
13) T	2,4-DB	15.567	15.073	197.9E6	192.1E6	757.153	756.496
14) T	DINOSEB	17.037	15.498	1355.5E6	1256.0E6	720.492	716.488
15) T	Picloram	16.808	16.728	2334.6E6	2200.3E6	737.952	731.857
16) T	DCPA	17.383	16.673	2229.2E6	2256.4E6	741.266	728.436

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020519\
Data File : PS003029.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Feb 2019 16:32
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
2/6/2019 4:17:58 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 06 00:13:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012819.M
Quant Title : 8080.M
QLast Update : Tue Jan 29 06:11:45 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

