

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020520\
 Data File : PS008736.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Feb 2020 17:14
 Operator : DD\AJ
 Sample : L1345-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 06 00:42:12 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020420.M
 Quant Title : 8080.M
 QLast Update : Wed Feb 05 02:36:36 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.250	6.945	628.6E6	299.1E6	292.892	322.694
Target Compounds							
1) T	Dalapon	0.000	2.593	0	194.7E6	N.D.	162.372
2) T	3,5-DICHL...	0.000	6.119	0	96854419	N.D.	70.394
7) T	MCPA	7.803	7.413	70104592	46559618	8.161	11.156 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020520\
Data File : PS008736.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Feb 2020 17:14
Operator : DD\AJ
Sample : L1345-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 06 00:42:12 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020420.M
Quant Title : 8080.M
QLast Update : Wed Feb 05 02:36:36 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 0.000 min
Exp R.T.: 3.051 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.593 min
Delta R.T.: -0.039 min
Response: 194735988
Conc: 162.37 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T.: 6.458 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.119 min
Delta R.T.: 0.031 min
Response: 96854419
Conc: 70.39 ng/ml



#4 2,4-DCAA


R.T.: 7.250 min
Delta R.T.: -0.002 min
Response: 628636131
Conc: 292.89 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 6.945 min
Delta R.T.: -0.001 min
Response: 299080555
Conc: 322.69 ng/ml



#7 MCPA

R.T.: 7.803 min
Delta R.T.: 0.030 min
Response: 70104592
Conc: 8.16 ug/ml



#7 MCPA

R.T.: 7.413 min
Delta R.T.: -0.027 min
Response: 46559618
Conc: 11.16 ug/ml