

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020520\
 Data File : PS008731.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Feb 2020 15:20
 Operator : DD\AJ
 Sample : L1004-18MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 BU-01-012920MSD

Manual Integrations
 APPROVED

mohammad
 2/7/2020 5:04:07 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 05 15:44:40 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020420.M
 Quant Title : 8080.M
 QLast Update : Wed Feb 05 02:36:36 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.251	6.946	752.9E6	352.6E6	350.781	380.430
Target Compounds							
1) T	Dalapon	3.049	2.632	1109.2E6	777.4E6	437.323	648.182 #
2) T	3,5-DICHL...	6.457	6.087	1475.1E6	681.5E6	522.298	495.296
5) T	DICAMBA	7.444	7.117	4109.4E6	2001.9E6	509.475	514.450
6) T	MCPD	7.613	7.213	339.0E6	168.0E6	55.156	57.245
7) T	MCPA	7.763	7.429	491.0E6	240.8E6	57.160	57.689
8) T	DICHLORPROP	8.140	7.746	1109.0E6	549.9E6	560.546	577.234
9) T	2,4-D	8.369	8.040	1329.1E6	724.3E6	621.064	634.501
10) T	Pentachlo...	8.683	8.474	11411.3E6	5408.1E6	401.398	410.044
11) T	2,4,5-TP ...	9.233	8.834	6542.9E6	3260.2E6	585.340	636.104m
12) T	2,4,5-T	9.517	9.215	5947.8E6	2885.2E6	540.410	558.328
13) T	2,4-DB	10.071	9.724	787.6E6	373.8E6	484.199	534.621
14) T	DINOSEB	11.237	10.059	2423.7E6	1063.1E6	287.212	289.426
15) T	Picloram	11.050	11.040	6500.6E6	3207.2E6	429.257	424.896
16) T	DCPA	11.502	10.991	5176.2E6	2493.6E6	369.233	383.381

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020520\
Data File : PS008731.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Feb 2020 15:20
Operator : DD\AJ
Sample : L1004-18MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
BU-01-012920MSD

Manual Integrations
APPROVED

mohammad
2/7/2020 5:04:07 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 15:44:40 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020420.M
Quant Title : 8080.M
QLast Update : Wed Feb 05 02:36:36 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

