

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020521\
 Data File : PS013951.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Feb 2021 17:04
 Operator : DD\AJ
 Sample : M1286-17MSD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 JJ_BH_03_0-1MSD

Manual Integrations
 APPROVED

Ankita
 2/9/2021 9:29:43 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 05 17:42:52 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012721.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 28 01:00:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.205	6.904	474.9E6	222.7E6	441.375	577.339 #
Target Compounds							
1) T	Dalapon	3.025	2.620	464.3E6	275.6E6	372.027	536.083 #
2) T	3,5-DICHL...	6.422	6.061	457.6E6	186.5E6	307.960	307.185
3) T	4-Nitroph...	7.039	6.519	226.4E6	99386090	416.580	395.213
5) T	DICAMBA	7.394	7.070	1346.7E6	595.1E6	313.194	361.139
6) T	MCPD	7.558	7.160	93993772	43022472	29.786	32.032
7) T	MCPA	7.707	7.373	131.3E6	78113028	29.017	40.295m#
8) T	DICHLORPROP	8.085	7.695	306.8E6	140.9E6	282.287	331.051
9) T	2,4-D	8.318	7.992	394.6E6	176.7E6	321.339	331.528
10) T	Pentachlo...	8.623	8.416	4392.0E6	1820.4E6	277.335	326.774
11) T	2,4,5-TP ...	9.171	8.776	1772.0E6	784.6E6	282.215	341.786
12) T	2,4,5-T	9.459	9.158	1786.4E6	727.9E6	277.338	326.912
13) T	2,4-DB	10.012	9.666	240.4E6	87165289	217.758	291.505 #
14) T	DINOSEB	11.163	9.994	245.2E6	61699680	48.383	37.767
15) T	Picloram	10.998	10.987	2617.9E6	1235.9E6	308.315	362.161
16) T	DCPA	11.431	10.919	2758.0E6	1301.2E6	341.179	434.342 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020521\
Data File : PS013951.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Feb 2021 17:04
Operator : DD\AJ
Sample : M1286-17MSD
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
JJ_BH_03_0-1MSD

Manual Integrations
APPROVED

Ankita
2/9/2021 9:29:43 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 17:42:52 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012721.M
Quant Title : 8080.M
QLast Update : Thu Jan 28 01:00:30 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

