

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020524\
 Data File : PS025755.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Feb 2024 03:33
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 06 08:48:49 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011124.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 11 16:45:24 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.348	6.756	948.3E6	486.2E6	636.692	633.250
Target Compounds							
1) T	Dalapon	2.475	2.529	1323.1E6	656.4E6	571.760	558.707
2) T	3,5-DICHL...	5.725	5.950	1375.9E6	687.9E6	601.263	590.822
3) T	4-Nitroph...	6.189	6.376	550.1E6	276.5E6	608.756	580.696
5) T	DICAMBA	6.488	6.914	3947.9E6	2044.4E6	618.665	614.309
6) T	MCPD	6.638	7.011	254.8E6	143.3E6	63.636	61.763
7) T	MCPA	6.759	7.216	357.9E6	204.9E6	63.241	59.132
8) T	DICHLORPROP	7.074	7.527	1004.9E6	547.7E6	609.802	633.592
9) T	2,4-D	7.274	7.810	1040.7E6	516.0E6	565.183	567.971
10) T	Pentachlo...	7.498	8.230	15097.6E6	7567.3E6	630.887	615.230
11) T	2,4,5-TP ...	8.020	8.589	5714.3E6	2901.3E6	614.850	614.595
12) T	2,4,5-T	8.278	8.959	5364.7E6	2317.4E6	549.452	530.209
13) T	2,4-DB	8.783	9.463	654.0E6	319.9E6	607.605	612.818
14) T	DINOSEB	9.828	9.787	3000.1E6	1261.1E6	437.487	401.139
15) T	Picloram	9.684	10.752	2018.8E6	1040.6E6	195.128	195.720
16) T	DCPA	10.094	10.706	7243.7E6	3282.3E6	597.224	564.510

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020524\
Data File : PS025755.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Feb 2024 03:33
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 06 08:48:49 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS011124.M
Quant Title : 8080.M
QLast Update : Thu Jan 11 16:45:24 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

