

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020624\
 Data File : PS025773.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Feb 2024 15:18
 Operator : AR\AJ
 Sample : HSTDICC1500
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 06 15:36:13 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020624.M
 Quant Title : 8080.M
 QLast Update : Tue Feb 06 15:36:04 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.346	6.753	1893.3E6	984.7E6	1341.546	1399.779
Target Compounds							
1) T	Dalapon	2.467	2.523	2992.0E6	1447.5E6	1264.872	1304.940
2) T	3,5-DICHL...	5.723	5.947	2717.8E6	1381.2E6	1261.105	1310.255
3) T	4-Nitroph...	6.185	6.372	1201.4E6	585.2E6	1294.355	1280.766
5) T	DICAMBA	6.486	6.912	7879.8E6	4093.8E6	1309.272	1348.932
6) T	MCPD	6.640	7.013	566.5E6	299.6E6	155.722	146.424
7) T	MCPA	6.763	7.218	763.4E6	423.2E6	141.800	135.849
8) T	DICHLORPROP	7.071	7.524	1995.3E6	1045.3E6	1269.489	1308.921
9) T	2,4-D	7.269	7.806	2188.5E6	1132.2E6	1231.438	1327.274
10) T	Pentachlo...	7.496	8.227	28818.6E6	14676.0E6	1261.568	1293.738
11) T	2,4,5-TP ...	8.017	8.586	11485.0E6	5783.0E6	1300.675	1337.959
12) T	2,4,5-T	8.273	8.955	12118.2E6	5672.4E6	1280.138	1364.328
13) T	2,4-DB	8.778	9.457	1601.2E6	670.3E6	1383.338	1513.133
14) T	DINOSEB	9.825	9.784	8456.3E6	3808.7E6	1258.281	1324.685
15) T	Picloram	9.659	10.736	16189.3E6	7674.8E6	1376.546	1427.263
16) T	DCPA	10.091	10.703	14976.4E6	6508.4E6	1277.428	1311.035

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020624\
Data File : PS025773.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Feb 2024 15:18
Operator : AR\AJ
Sample : HSTDICC1500
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 06 15:36:13 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020624.M
Quant Title : 8080.M
QLast Update : Tue Feb 06 15:36:04 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

