

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020624\
 Data File : PS025776.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Feb 2024 16:30
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 06 16:50:53 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020624.M
 Quant Title : 8080.M
 QLast Update : Tue Feb 06 15:37:16 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.346	6.754	989.6E6	504.1E6	701.208	716.656
Target Compounds							
1) T	Dalapon	2.466	2.522	1525.6E6	725.2E6	644.960	653.796
2) T	3,5-DICHL...	5.723	5.948	1433.3E6	707.0E6	665.062	670.660
3) T	4-Nitroph...	6.186	6.374	599.4E6	294.4E6	645.781	644.278
5) T	DICAMBA	6.486	6.912	4085.8E6	2084.3E6	678.882	686.775
6) T	MCPD	6.636	7.009	262.8E6	143.1E6	72.244	69.955
7) T	MCPA	6.757	7.214	371.3E6	210.1E6	68.965	67.451
8) T	DICHLORPROP	7.071	7.524	1049.3E6	557.8E6	667.609	698.456
9) T	2,4-D	7.271	7.807	1157.4E6	575.4E6	651.263	674.589
10) T	Pentachlo...	7.495	8.226	15745.2E6	7906.6E6	689.263	696.990
11) T	2,4,5-TP ...	8.017	8.586	6097.0E6	3000.9E6	690.487	694.293
12) T	2,4,5-T	8.274	8.955	6478.8E6	2879.5E6	684.406	692.575
13) T	2,4-DB	8.780	9.458	789.3E6	334.8E6	681.922	755.919
14) T	DINOSEB	9.825	9.784	4490.3E6	1927.5E6	668.143	670.405
15) T	Picloram	9.663	10.738	8120.6E6	3752.8E6	690.477	697.900
16) T	DCPA	10.091	10.703	8147.3E6	3478.3E6	694.928	700.659

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020624\
Data File : PS025776.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Feb 2024 16:30
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 06 16:50:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020624.M
Quant Title : 8080.M
QLast Update : Tue Feb 06 15:37:16 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

