

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020922\
 Data File : PS018308.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Feb 2022 20:00
 Operator : AJ\MA
 Sample : N1432-16MS
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 SP-6MS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 02/10/2022
 Supervised By :Ankita Jodhani 02/10/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 10 00:12:53 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020822.M
 Quant Title : 8080.M
 QLast Update : Wed Feb 09 15:57:26 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.025	6.638	593.1E6	260.9E6	377.008	415.930
Target Compounds							
1) T	Dalapon	2.887	2.407	304.1E6	214.1E6	121.039	267.193m#
2) T	3,5-DICHL...	6.271	5.859	444.6E6	153.2E6	198.765	161.388
3) T	4-Nitroph...	6.862	6.282	191.2E6	66822999	216.680	164.995
5) T	DICAMBA	7.210	6.790	1260.6E6	522.5E6	199.566m	192.610m
6) T	MCP	7.373	6.878	79135895	37870777	17.195	17.464
7) T	MCPA	7.519	7.084	120.5E6	68268025	18.491	22.119
8) T	DICHLORPROP	7.896	7.402	312.4E6	136.3E6	197.999	196.625
9) T	2,4-D	8.128	7.693	413.4E6	176.0E6	224.519	212.337
10) T	Pentachlo...	8.418	8.098	3590.8E6	1487.0E6	169.924	172.265
11) T	2,4,5-TP ...	8.967	8.460	1708.8E6	735.2E6	200.009	198.489
12) T	2,4,5-T	9.253	8.842	1904.1E6	701.7E6	215.160	198.190
13) T	2,4-DB	9.803	9.344	236.5E6	81681693	134.403	171.578 #
14) T	DINOSEB	10.938	9.662	365.3E6	97623415	54.708	38.684 #
15) T	Picloram	10.777	10.645	2355.8E6	1072.7E6	206.201	214.023
16) T	DCPA	11.208	10.584	2159.0E6	1110.3E6	195.874	243.953

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS020922\
Data File : PS018308.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Feb 2022 20:00
Operator : AJ\MA
Sample : N1432-16MS
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
SP-6MS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 02/10/2022
Supervised By : Ankita Jodhani 02/10/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 10 00:12:53 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020822.M
Quant Title : 8080.M
QLast Update : Wed Feb 09 15:57:26 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

