

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021023\
 Data File : PS022128.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Feb 2023 02:25
 Operator : AR\AJ
 Sample : 01510-16MSD
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 JPP-59-020823MSD

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 02/13/2023
 Supervised By : Ankita Jodhani 02/13/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 11 04:09:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020223.M
 Quant Title : 8080.M
 QLast Update : Fri Feb 03 10:03:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.299	6.726	1184.7E6	1240.9E6	519.743	613.025
Target Compounds							
1) T	Dalapon	2.417	2.492	977.1E6	1443.7E6	314.708m	463.962m#
2) T	3,5-DICHL...	5.680	5.925	1846.3E6	1606.8E6	536.140	537.719
5) T	DICAMBA	6.436	6.884	4511.1E6	4364.4E6	466.379	461.958
6) T	MCP	6.587	6.982	413.4E6	377.3E6	57.415	50.552
7) T	MCPA	6.708	7.191	475.1E6	541.3E6	47.443	53.891m
8) T	DICHLORPROP	7.019	7.505	1197.6E6	1222.1E6	484.396	510.809
9) T	2,4-D	7.220	7.795	1495.0E6	1521.9E6	497.247	539.722
10) T	Pentachlo...	7.441	8.214	18034.8E6	17023.6E6	546.189	535.832
11) T	2,4,5-TP ...	7.968	8.577	6991.1E6	6902.9E6	524.729	524.117
12) T	2,4,5-T	8.226	8.953	7018.2E6	6345.5E6	489.016	525.083
13) T	2,4-DB	8.734	9.458	1017.2E6	848.6E6	462.635	502.545
14) T	DINOSEB	9.777	9.781	4781.0E6	4626.0E6	464.891	494.904
15) T	Picloram	9.617	10.751	8570.3E6	8024.7E6	460.185	442.159
16) T	DCPA	10.045	10.707	9207.4E6	7914.4E6	541.896	514.252

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021023\
Data File : PS022128.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Feb 2023 02:25
Operator : AR\AJ
Sample : 01510-16MSD
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
JPP-59-020823MSD

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 02/13/2023
Supervised By : Ankita Jodhani 02/13/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 11 04:09:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020223.M
Quant Title : 8080.M
QLast Update : Fri Feb 03 10:03:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

