

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021125\
 Data File : PS029118.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Feb 2025 17:56
 Operator : AR\AJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 12 01:06:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021125.M
 Quant Title : 8080.M
 QLast Update : Wed Feb 12 01:05:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.187	7.665	688.8E6	253.1E6	260.349	248.013
Target Compounds							
1) T	Dalapon	2.607	2.656	637.7E6	376.8E6	207.145	198.486
2) T	3,5-DICHL...	6.365	6.633	904.5E6	327.5E6	237.115	219.637
3) T	4-Nitroph...	6.985	7.195	357.2E6	181.0E6	217.559	216.114
5) T	DICAMBA	7.371	7.861	2662.7E6	1134.9E6	226.207	208.306
6) T	MCP	7.549	7.962	122.9E6	46510552	17.524	18.760
7) T	MCPA	7.697	8.202	177.9E6	66885235	18.485	19.096
8) T	DICHLORPROP	8.073	8.573	735.9E6	308.2E6	242.594	230.623
9) T	2,4-D	8.303	8.898	779.3E6	311.1E6	240.704	222.548
10) T	Pentachlo...	8.598	9.419	11502.2E6	5239.6E6	238.382	219.299
11) T	2,4,5-TP ...	9.173	9.797	4474.4E6	2100.7E6	237.526	218.016
12) T	2,4,5-T	9.464	10.214	4468.3E6	1998.9E6	236.559	219.930
13) T	2,4-DB	10.034	10.778	879.5E6	203.4E6	249.293	219.614
14) T	DINOSEB	11.235	11.155	3250.2E6	1139.8E6	234.520	213.961
15) T	Picloram	11.047	12.235	7142.6E6	2733.0E6	229.648	198.993
16) T	DCPA	11.530	12.192	6856.2E6	2641.9E6	242.903	216.565

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021125\
Data File : PS029118.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Feb 2025 17:56
Operator : AR\AJ
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC200

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 12 01:06:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021125.M
Quant Title : 8080.M
QLast Update : Wed Feb 12 01:05:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

