

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021320\
 Data File : PS008833.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Feb 2020 14:31
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 14 02:07:29 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020420.M
 Quant Title : 8080.M
 QLast Update : Wed Feb 05 02:36:36 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.249	6.943	1732.7E6	759.0E6	807.291	818.894
Target Compounds							
1) T	Dalapon	3.049	2.631	1982.7E6	960.0E6	781.727	800.455
2) T	3,5-DICHL...	6.456	6.085	2105.7E6	1042.9E6	745.607	757.990
3) T	4-Nitroph...	7.077	6.543	787.3E6	417.9E6	725.861	734.965
5) T	DICAMBA	7.442	7.114	6201.9E6	3039.1E6	768.903	780.977
6) T	MCPD	7.613	7.212	483.9E6	231.6E6	78.732	78.916
7) T	MCPA	7.764	7.429	650.1E6	318.1E6	75.679	76.229
8) T	DICHLORPROP	8.137	7.742	1501.1E6	727.7E6	758.706	763.783
9) T	2,4-D	8.366	8.037	1593.0E6	873.0E6	744.371	764.799
10) T	Pentachlo...	8.680	8.471	21651.1E6	10247.1E6	761.592	776.937
11) T	2,4,5-TP ...	9.230	8.832	8497.3E6	3995.3E6	760.185	779.544
12) T	2,4,5-T	9.514	9.211	8284.7E6	4016.4E6	752.741	777.234
13) T	2,4-DB	10.067	9.719	1200.1E6	529.0E6	737.790	756.566
14) T	DINOSEB	11.234	10.056	6305.6E6	2865.2E6	747.234	780.057
15) T	Picloram	11.047	11.035	10847.2E6	5534.4E6	716.281	733.220
16) T	DCPA	11.498	10.986	10471.6E6	5061.4E6	746.960	778.180

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021320\
Data File : PS008833.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Feb 2020 14:31
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 02:07:29 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020420.M
Quant Title : 8080.M
QLast Update : Wed Feb 05 02:36:36 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

