

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021320\
 Data File : PS008850.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Feb 2020 22:53
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 2/14/2020 11:59:15 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 14 02:10:38 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020420.M
 Quant Title : 8080.M
 QLast Update : Wed Feb 05 02:36:36 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.247	6.942	1689.7E6	749.5E6	787.252	808.726
Target Compounds							
1) T	Dalapon	3.048	2.630	1861.1E6	888.9E6	733.793	741.168
2) T	3,5-DICHL...	6.455	6.085	2058.6E6	1024.2E6	728.911	744.372
3) T	4-Nitroph...	7.075	6.542	770.9E6	410.5E6	710.671	721.951
5) T	DICAMBA	7.441	7.113	6093.3E6	2985.1E6	755.445	767.116
6) T	MCPD	7.612	7.211	476.4E6	228.4E6	77.507	77.830
7) T	MCPA	7.761	7.428	644.3E6	312.3E6	74.999	74.832m
8) T	DICHLORPROP	8.135	7.741	1468.9E6	715.6E6	742.451	751.127
9) T	2,4-D	8.363	8.036	1554.4E6	854.9E6	726.324	748.928
10) T	Pentachlo...	8.677	8.469	21141.3E6	10037.5E6	743.659	761.048
11) T	2,4,5-TP ...	9.228	8.831	8306.2E6	3910.5E6	743.087	762.988
12) T	2,4,5-T	9.512	9.210	8078.0E6	3924.8E6	733.960	759.506
13) T	2,4-DB	10.065	9.718	1176.7E6	523.2E6	723.436	748.217
14) T	DINOSEB	11.232	10.055	6067.9E6	2741.5E6	719.065	746.384
15) T	Picloram	11.044	11.034	9873.7E6	5080.1E6	651.994	673.028
16) T	DCPA	11.496	10.985	10252.6E6	4990.3E6	731.345	767.238

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021320\
Data File : PS008850.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Feb 2020 22:53
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
2/14/2020 11:59:15 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 02:10:38 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020420.M
Quant Title : 8080.M
QLast Update : Wed Feb 05 02:36:36 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

