

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021325\
 Data File : PS029155.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Feb 2025 16:36
 Operator : AR\AJ
 Sample : PB166710BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB166710BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 14 01:19:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021225.M
 Quant Title : 8080.M
 QLast Update : Wed Feb 12 09:17:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.190	7.663	1636.4E6	586.5E6	592.074	554.126
Target Compounds							
1) T	Dalapon	2.613	2.660	1403.6E6	858.2E6	443.749	453.302
2) T	3,5-DICHL...	6.368	6.631	2197.1E6	813.0E6	554.581	531.409
3) T	4-Nitroph...	6.988	7.193	942.2E6	456.7E6	552.594	526.555
5) T	DICAMBA	7.374	7.858	6842.1E6	2967.3E6	568.189	536.908
6) T	MCPD	7.555	7.962	379.4E6	136.8E6	54.760	55.435
7) T	MCPA	7.703	8.202	526.2E6	190.1E6	54.973	54.348
8) T	DICHLORPROP	8.076	8.569	1789.7E6	733.1E6	566.523	532.352
9) T	2,4-D	8.306	8.895	1888.2E6	765.4E6	561.284	534.270
10) T	Pentachlo...	8.602	9.416	28847.5E6	13390.4E6	611.773	555.168
11) T	2,4,5-TP ...	9.176	9.793	11121.8E6	5367.5E6	573.575	549.487
12) T	2,4,5-T	9.467	10.209	11076.4E6	5014.8E6	571.842	541.421
13) T	2,4-DB	10.037	10.773	2026.9E6	503.6E6	545.417	530.109
14) T	DINOSEB	11.238	11.150	9065.2E6	3357.8E6	625.161	608.886
15) T	Picloram	11.051	12.231	16874.0E6	6415.7E6	528.096	466.636
16) T	DCPA	11.534	12.187	16981.9E6	6815.2E6	585.658	552.516

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021325\
Data File : PS029155.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Feb 2025 16:36
Operator : AR\AJ
Sample : PB166710BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB166710BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 01:19:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021225.M
Quant Title : 8080.M
QLast Update : Wed Feb 12 09:17:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

