

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021424\
 Data File : PS025936.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Feb 2024 13:17
 Operator : AR\AJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 14 15:03:26 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021424.M
 Quant Title : 8080.M
 QLast Update : Wed Feb 14 14:47:28 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.276	7.811	619.6E6	197.1E6	228.577	222.328
Target Compounds							
1) T	Dalapon	2.680	2.730	985.4E6	344.5E6	214.806	207.183
2) T	3,5-DICHL...	6.442	6.758	836.5E6	276.1E6	210.783	203.623
3) T	4-Nitroph...	7.077	7.339	324.7E6	97591799	202.241	199.948
5) T	DICAMBA	7.466	8.013	2270.8E6	769.3E6	203.077	197.700
6) T	MCPD	7.646	8.110	142.0E6	50366818	18.140	18.808
7) T	MCPA	7.797	8.355	223.9E6	77387082	19.604	20.005
8) T	DICHLORPROP	8.181	8.729	613.3E6	202.4E6	212.971	207.633
9) T	2,4-D	8.413	9.060	634.2E6	191.5E6	210.634	200.554
10) T	Pentachlo...	8.713	9.591	8660.0E6	3004.7E6	208.634	205.140
11) T	2,4,5-TP ...	9.298	9.967	3295.3E6	1125.9E6	207.345	204.399
12) T	2,4,5-T	9.593	10.388	3125.3E6	987.7E6	202.523	196.924
13) T	2,4-DB	10.173	10.955	387.4E6	108.6E6	197.595	197.218
14) T	DINOSEB	11.393	11.334	2270.4E6	719.6E6	201.817	198.849
15) T	Picloram	11.197	12.436	3348.8E6	933.1E6	188.204	177.857
16) T	DCPA	11.686	12.380	4186.4E6	1323.2E6	208.724	203.284

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021424\
Data File : PS025936.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Feb 2024 13:17
Operator : AR\AJ
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC200

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 15:03:26 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS021424.M
Quant Title : 8080.M
QLast Update : Wed Feb 14 14:47:28 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

