

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021519\
 Data File : PS003218.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Feb 2019 00:58
 Operator : EI\SJ
 Sample : K1484-04
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 M-2-S

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 16 01:38:44 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012819.M
 Quant Title : 8080.M
 QLast Update : Tue Jan 29 06:11:45 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.285	10.784	150.8E6	167.2E6	397.141	391.097
Target Compounds							
6) T	MCP	11.967	0.000	13005873	0	10.855	N.D. #
15) T	Picloram	0.000	16.711	0	7605196	N.D.	2.530 #
16) T	DCPA	0.000	16.711	0	7605196	N.D.	2.455 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021519\
Data File : PS003218.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Feb 2019 00:58
Operator : EI\SJ
Sample : K1484-04
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
M-2-S

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 16 01:38:44 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS012819.M
Quant Title : 8080.M
QLast Update : Tue Jan 29 06:11:45 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 11.285 min
Delta R.T.: -0.033 min
Response: 150771078
Conc: 397.14 ng/ml

Instrument :
ECD_S
ClientSampleId :
M-2-S




#4 2,4-DCAA

R.T.: 10.784 min
Delta R.T.: -0.033 min
Response: 167211779
Conc: 391.10 ng/ml




#6 MCP

R.T.: 11.967 min
Delta R.T.: -0.002 min
Response: 13005873
Conc: 10.86 ug/ml



#6 MCP


R.T.: 0.000 min
Exp R.T.: 11.330 min
Response: 0
Conc: N.D.



#15 Picloram


R.T.: 0.000 min
Exp R.T.: 16.806 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
M-2-S




#15 Picloram

R.T.: 16.711 min
Delta R.T.: -0.013 min
Response: 7605196
Conc: 2.53 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 17.382 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 16.711 min
Delta R.T.: 0.041 min
Response: 7605196
Conc: 2.46 ng/ml