

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021523\
 Data File : PS022142.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Feb 2023 19:17
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 14 20:27:26 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020223.M
 Quant Title : 8080.M
 QLast Update : Fri Feb 03 10:03:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.306	6.729	1163.3E6	1121.8E6	510.344	554.168
Target Compounds							
1) T	Dalapon	2.445	0.000	2333873	0	<MDL	N.D. #
2) T	3,5-DICHL...	5.689	0.000	2179690	0	<MDL	N.D. #
3) T	4-Nitroph...	6.174	0.000	4996755	0	3.713	N.D. #
5) T	DICAMBA	0.000	6.943f	0	13109734	N.D.	1.388
6) T	MCP	0.000	7.033f	0	6807890	N.D.	<MDL
7) T	MCPA	6.702	7.133f	807329	37356226	<MDL	3.719 #
9) T	2,4-D	7.288f	0.000	29935964	0	9.957	N.D. #
10) T	Pentachlo...	7.465	8.225	7113050	6451812	<MDL	<MDL #
11) T	2,4,5-TP ...	7.950	8.607	336425	3129365	<MDL	<MDL #
12) T	2,4,5-T	8.191f	0.000	148432	0	<MDL	N.D. #
14) T	DINOSEB	9.789	9.794	2104353	2607794	<MDL	<MDL #
15) T	Picloram	9.645	10.712f	8163072	7862161	<MDL	<MDL #
16) T	DCPA	10.053	10.712	4503206	7862161	<MDL	<MDL #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021523\
Data File : PS022142.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Feb 2023 19:17
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 20:27:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020223.M
Quant Title : 8080.M
QLast Update : Fri Feb 03 10:03:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#3 4-Nitrophenol


R.T.: 6.174 min
Delta R.T.: 0.028 min
Response: 4996755
Conc: 3.71 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T. : 6.354 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 6.306 min
Delta R.T.: 0.004 min
Response: 1163258997
Conc: 510.34 ng/ml



#4 2,4-DCAA


R.T.: 6.729 min
Delta R.T.: -0.001 min
Response: 1121769781
Conc: 554.17 ng/ml



#5 DICAMBA


R.T.: 0.000 min
Exp R.T.: 6.438 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 6.943 min
Delta R.T.: 0.054 min
Response: 13109734
Conc: 1.39 ng/ml




#7 MCPA

R.T.: 6.702 min
Delta R.T.: -0.011 min
Response: 807329
Conc: N.D.



#7 MCPA


R.T.: 7.133 min
Delta R.T.: -0.065 min
Response: 37356226
Conc: 3.72 ug/ml



#9 2,4-D

R.T.: 7.288 min
Delta R.T.: 0.064 min
Response: 29935964
Conc: 9.96 ng/ml

Instrument :
ECD_S
ClientSampleId :



#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 7.799 min
Response: 0
Conc: N.D.