

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021720\
 Data File : PS008869.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Feb 2020 15:34
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 17 15:52:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020420.M
 Quant Title : 8080.M
 QLast Update : Wed Feb 05 02:36:36 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.247	6.942	1734.9E6	758.7E6	808.323	818.585
Target Compounds							
1) T	Dalapon	3.048	2.632	1957.8E6	920.3E6	771.936	767.345
2) T	3,5-DICHL...	6.455	6.085	2120.1E6	1037.4E6	750.696	753.979
3) T	4-Nitroph...	7.076	6.543	790.5E6	417.2E6	728.808	733.716
5) T	DICAMBA	7.441	7.113	6270.4E6	3027.4E6	777.395	777.981
6) T	MCPD	7.612	7.211	485.3E6	232.8E6	78.960	79.321
7) T	MCPA	7.762	7.428	663.6E6	321.4E6	77.253	77.009
8) T	DICHLORPROP	8.135	7.741	1509.5E6	726.2E6	762.962	762.205
9) T	2,4-D	8.363	8.035	1629.3E6	871.3E6	761.338	763.342
10) T	Pentachlo...	8.678	8.470	21576.3E6	10087.0E6	758.957	764.801
11) T	2,4,5-TP ...	9.227	8.830	8536.5E6	3973.2E6	763.687	775.239
12) T	2,4,5-T	9.511	9.210	8324.1E6	4012.1E6	756.323	776.393
13) T	2,4-DB	10.065	9.718	1191.8E6	530.9E6	732.711	759.197
14) T	DINOSEB	11.232	10.054	6235.5E6	2814.3E6	738.925	766.200
15) T	Picloram	11.044	11.034	8907.4E6	4430.3E6	588.189	586.943
16) T	DCPA	11.496	10.985	10429.3E6	5029.3E6	743.948	773.233

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS021720\
Data File : PS008869.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Feb 2020 15:34
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 17 15:52:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS020420.M
Quant Title : 8080.M
QLast Update : Wed Feb 05 02:36:36 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

